

Activator first off with Filter  
Base first off with Filter



Diversey Europe Operations BV

Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 27/9/12

<b>Product Name</b>	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26
<b>Product Code</b>	7515828
<b>Batch Number</b>	FMP12 272 32942
<b>Production Date</b>	09/2012
<b>Expiration Date</b>	09/2013

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear colourless liquid
pH (100% 20°C)	JDM001	9.0	12.5	11.6
Specific Gravity (20°C)	JDM004	1.004	1.020	1.009
Potential Chlorine Dioxide (ppm)		100	180	168.75 ppm

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Bi-Spore Activator	Version : 01	Date of issuing : October 1 <sup>st</sup> 2012
------------------------	--------------	--

Activator First off With Filter  
Base First off With Filter



Diversey Europe Operations BV

Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 29.9.12

<b>Product Name</b>	CLEARKLENS BI-SPORE BASE SOLUTION VH26
<b>Product Code</b>	7515828
<b>Batch Number</b>	EMP12272 32942
<b>Production Date</b>	09/2012
<b>Expiration Date</b>	09/2013

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear colourless liquid
pH (100% 20°C)	JDM001	1.5	2.5	1.7
Specific Gravity (20°C)	JDM004	1.010	1.030	1.017

On behalf of Diversey site Quality Manager	Name :	Makine
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Bi-Spore Base	Version : 01	Date of issuing : September 27 <sup>th</sup> 2012
-------------------	--------------	---

# CERTIFICATE OF IRRADIATION

Number: **B**

0546330 00

Page 1

FLE001

Flexible Medical Packaging Ltd  
Unit 8  
White Cross Industrial Estate  
Hightown  
Lancaster

LA1 4XS  
F019134  
Cust. Ref: 13/09/12  
Date Rec'd: 18/09/12  
Date:



## Synergy Health plc

Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@synergyhealthplc.com  
Website: www.synergyhealthplc.com

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010113 B212090481	JD-BOTTLES/CAPS/BAGS 25-40kGy	24	8 X B/No. W032931 = FMPI2272.32942. 8 X B/No. W032933 (25.10.12) 8 X B/No. W032932  CUSTOMER BATCH/CODE No.s AS ORDER No. F019134 DATED 12/9/12
	Despatch Text:  6PLTS		
	Total	24	Last Page 1 of 1

All in accordance with current  
Technical Agreement

This is to certify that the above items have been irradiated as  
specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



# Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,  
Hoeford Point, Barwell Lane, Gosport  
Hampshire PO13 0AU England  
Telephone: +44(0)1329 226600  
+44(0)1329 832511  
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk  
www.wickhamlabs.co.uk

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 03 Oct 2012  
**Date Tested:** 09 Oct 2012  
**Date Test Completed:** 23 Oct 2012  
**Purchase Order:** PO19227

## CERTIFICATE OF ANALYSIS

**Laboratory Reference Number:** 0006988/2  
**Test Required:** Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP  
**Date Received:** 03/10/2012  
**Test Article:** Bi-Spore Activator  
**Sample Code:** 415168  
**Batch Ref:** FMP12272 32942  
**Qty Received:** 20 x 100mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM04

### Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 24 Oct 2012 15:38:19

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0006988

Certificate of Analysis - OSMM

Print Number: P0006386

Page 1 of 1





**Wickham Laboratories**  
contract analytical services

Wickham Laboratories Limited,  
Hoeford Point, Barwell Lane, Gosport  
Hampshire PO13 0AU England  
Telephone: +44(0)1329 226600  
+44(0)1329 832511  
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk  
www.wickhamlabs.co.uk

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 03 Oct 2012  
**Date Tested:** 09 Oct 2012  
**Date Test Completed:** 23 Oct 2012  
**Purchase Order:** PO19227

### CERTIFICATE OF ANALYSIS

**Laboratory Reference Number:** 0006988/1  
**Test Required:** Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP  
**Date Received:** 03/10/2012  
**Test Article:** Bi-Spore Base  
**Sample Code:** 415169  
**Batch Ref:** FMP12272 32942  
**Qty Received:** 20 x 100mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM05

### Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

*Suz Wood*

Date: 24 Oct 2012 15:38:19

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0006988

Certificate of Analysis - OSMM

Print Number: P0006386

Page 1 of 1