

Activator first off with filter
Base first off with filter



Diversey Europe Operations BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 31/10/12

Product Name	CLEARLENDS BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7514852
Batch Number	FMP12277 32944
Production Date	10/2012
Expiration Date	10/2013

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless liquid
pH (100% 20°C)	JDM001	9.0	12.5	11.6
Specific Gravity (20°C)	JDM004	1.004	1.020	1.009
Potential Chlorine Dioxide (ppm)		100	180	168.75ppm

On behalf of Diversey site Quality Manager	Name :	C. Rascoe
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Bi-Spore Activator	Version : 01	Date of issuing : October 1 st 2012
------------------------	--------------	--

Activator first off with filter
Base first off with filter



Diversey Europe Operations BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date 3/10/12

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Product Code	7514852
Batch Number	FMP12277 32944
Production Date	10/2012
Expiration Date	10/2013

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear colourless liquid
pH (100% 20°C)	JDM001	1.5	2.5	1.7
Specific Gravity (20°C)	JDM004	1.010	1.030	1.017

On behalf of Diversey site Quality Manager	Name :	C. Pascoe
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Bi-Spore Base	Version : 01	Date of issuing : September 27 th 2012
-------------------	--------------	---



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 12 Oct 2012
Date Tested: 16 Oct 2012
Date Test Completed: 30 Oct 2012
Purchase Order: PO19285

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0007188/2
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 12/10/2012
Test Article: Bi-Spore Activator
Sample Code: 415168
Batch Ref: FMP12277 32944
Qty Received: 20 x 100mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50mL
Product Standard Data Sheet	FG047/psd	FM04

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 01 Nov 2012 12:29:33

Mrs C Moore, CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0007188

Certificate of Analysis - OSMM

Print Number: P0006614 Page 1 of 1



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 12 Oct 2012
Date Tested: 16 Oct 2012
Date Test Completed: 30 Oct 2012
Purchase Order: PO19285

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0007188/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 12/10/2012
Test Article: BI-Spore Base
Sample Code: 415169
Batch Ref: FMP12277 32944
Qty Received: 20 x 100mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50mL
Product Standard Data Sheet	FG047/psd	FM05

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 01 Nov 2012 12:29:32

Mrs C Moore, CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0007188

Certificate of Analysis - OSMM

Print Number: P0006614 Page 1 of 1

CERTIFICATE OF IRRADIATION

Number: **B**

0546330 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster

Synergy Health plc

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com



LA1 4XS

P019134

Cust. Ref: 13/09/12

Date Rec'd 18/09/12

Date: 18/09/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010113 B212090481	JD-BOTTLES/CAPS/BAGS 25-40kGy	24	8 X B/No.W032931 8 X B/No.W032933 8 X B/No.W032932 FMP12277 32944 (EQ) 25.10.12. CUSTOMER BATCH/CODE No.s AS ORDER No.P019134 DATED 12/9/12
	Despatch Text: 6PLTS		
	Total	24	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature: *[Signature]*

For and on behalf of SYNERGY HEALTH PLC