



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 12-Mar-10
Date Tested : 24-Mar-10
Date Completed : 07-Apr-10
Purchase Order No. : PO15208

Test Article : Bi Spore Activator
Sample Code : Works Order No: 27562
Batch Reference : 00541
Laboratory Reference : LR-00940324 **Article Quantity** : 20
Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/2
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/2
Product Standard Data Sheet	*	MQ002/01	FM04

Day Book Ref: 800013
Volume Tested: All
Test carried out according to current Pharmacopoeias.

Mr K A Barker, CBiol., MSB
Business Manager

WXB

Date 14 Apr 2010

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

GLP

Compliant
Laboratory

Sample Booking Number: CN\00147842
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 2 of 2

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 12-Mar-10
Date Tested : 24-Mar-10
Date Completed : 07-Apr-10
Purchase Order No. : PO15208

Test Article : Bi Spore Base
Sample Code : Works Order No: 27562
Batch Reference : 00540
Laboratory Reference : LR-00940323 Article Quantity : 20
Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/2
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/2
Product Standard Data Sheet	*	MQ002/01	FM05

Day Book Ref: 800013
Volume Tested: All
Test carried out according to current Pharmacopoeias.

Mr K A Barker, CBiol., MSB
Business Manager

Date 14 Apr 2010

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

GLP

Compliant
Laboratory

Sample Booking Number: CN\00147842
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 2

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P015127
Date Rec'd 18/02/10
Date: 26/02/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010113 B210020442	JD-BOTTLES/CAPS/BAGS 25-40kGy	12	B/No.W027558 CUSTOMERS BATCH/CODE No.s AS ORDER No.P015127 DATED 17/2/10
Despatch Text: 2PLTS			
Total		12	Last Page 1 of 1
		All in accordance with current technical agreement	
		This is to certify that the above items have been irradiated as specified above	
		Authorising Signature: <i>J. Gifford</i> For and on behalf of ISOTRON LIMITED	

JohnsonDiversey
Clean is just the beginning



Contact address:

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date:

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7514852
Batch Number	NO 27562
Production Date	03/2010
Expiry Date	03/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
pH (100% 20°C)	JDM001	11.5 - 13.5	11.5
Specific Gravity (20°C)	JDM004	1.015 - 1.035	1.016

Signature / name

On behalf of Manufacturing Location



Removal of oil and dirt
Base last off with filter

JohnsonDiversey
Clean is just the beginning



Contact address:

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date:

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Product Code	7514852
Batch Number	WO 27562
Production Date	03/2010
Expiry Date	03/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Blue Liquid	clear blue liquid
pH (100% 20°C)	JDM001	1.5 - 3.5	1.9
Specific Gravity (20°C)	JDM004	1.020 - 1.060	1.030

Signature / name

On behalf of Manufacturing Location



Activator not off with filter
Base last off with filter



Contact address:

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date:

Product Name	CLEARKLENS BI-SPORE COMBINED SOLUTION VH26
Product Code	7514852
Batch Number	W0 27562
Production Date	03/2010
Expiry Date	03/2012

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Yellow Liquid	clear yellow liquid
pH (100% 20°C)	JDM001	2.5 - 5.5	2.8
Specific Gravity (20°C)	JDM004	1.000 - 1.010	1.001
Potential Chlorine Dioxide	Titration	230 - 280ppm	256.5 ppm

Signature / name

On behalf of Manufacturing Location

