



**Contact**

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 18/3/11

|                        |                                |
|------------------------|--------------------------------|
| <b>Product Name</b>    | CLEARLENSE CLEANSINALD SS VH9S |
| <b>Product Code</b>    | 7516431                        |
| <b>Batch Number</b>    | W029589                        |
| <b>Production Date</b> | 03/2011                        |
| <b>Expiry Date</b>     | 09/2012                        |

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

| Test                    | Test Method | Limits                       | Results                      |
|-------------------------|-------------|------------------------------|------------------------------|
| Appearance              | Visual      | Clear Slightly Yellow Liquid | clear slightly yellow liquid |
| pH (100% 20°C)          | JDM001      | 9.0 - 11.7                   | 10.9                         |
| Specific Gravity (20°C) | JDM004      | 0.990 - 1.010                | 0.997                        |

"SIGNATURE" C Pascoe

Name Caroline Pascoe  
Function Name Pharmacist Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

# CERTIFICATE OF IRRADIATION

Number: **B**

0480565 00

FLE001

Page 1

Flexible Medical Packaging Ltd  
Unit 8  
White Cross Industrial Estate  
Hightown  
Lancaster



## Isotron Limited

Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@isotron.com  
Website: www.isotron.com

LA1 4XS

P016641

Cust. Ref:

Date Rec'd 3/03/11

Date: 9/03/11

| ITEM CODE<br>ISOTRON BATCH  | ITEM SPECIFICATION                | QTY | ADDITIONAL DETAILS                                                                                                                              |
|-----------------------------|-----------------------------------|-----|-------------------------------------------------------------------------------------------------------------------------------------------------|
| B2FLE0010129<br>B211030205  | C/SINALD SS 900ml BOTTLE 25-40kGy | 11  | B/No.W029588 - <del>W029589</del><br><br>ACTUAL DOSE....33.9 - 34.2....kGy<br><br>CUSTOMERS BATCH/CODE No.s AS ORDER<br>No.P016641 DATED 2/3/11 |
| Despatch Text:<br><br>2PLTS |                                   |     |                                                                                                                                                 |
| Total                       |                                   | 11  | Last Page 1 of 1                                                                                                                                |

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



**Wickham Laboratories**  
contract analytical services

Wickham Laboratories Limited,  
Winchester Road, Wickham, Fareham,  
Hampshire PO17 5EU England  
Telephone: +44(0)1329 832511  
Fax: +44(0)1329 834262

mail@wickhamlabs.co.uk  
www.wickhamlabs.co.uk

## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

Date Received: 22 Mar 11  
Date Tested: 24 Mar 11  
Date Completed: 07 Apr 11  
Purchase Order: 16739

Test Article: ClearKlens Cleansinald SS  
Sample Code: 2505  
Batch Ref: WO29589  
Qty Received: 20 x 900 mL Bottles  
Laboratory Reference: 0001304/1  
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

| Test                                                           | Method Item | Result                 |
|----------------------------------------------------------------|-------------|------------------------|
| Sterility Test by Membrane Filtration (Steritest) Method       | MM107/00    | Pass                   |
| Growth in Tryptone Soya Broth at 20-25°C after 14 days         | MM107/01    | No growth in one broth |
| Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days | MM107/02    | No growth in one broth |
| Product Standard Data Sheet                                    | MM107/03    | FM06                   |
| Volume Tested                                                  | MM107/05    | 20 x 50 mL             |

### External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 08 Apr 11 15:50:32

Mrs S.L. Wood, BSc (Hons), CBIOL, MSB  
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0001304  
Lancaster

Certificate of Analysis - OSMM

Print Number

P0003187

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*