



Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 8/8/11

Product Name	CLEARKLENS TEGO 2000SC VH25
Product Code	7514341
Batch Number	W030188
Production Date	07/2011
Expiry Date	07/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH 100% (20°C)	JMD001	7.5 - 8.5	7.5
Specific Gravity (20°C)	JMD004	0.993 - 1.003	0.997

"SIGNATURE"

Name

Caroline Pascoe

Function Name

Pharmaceutical Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P017284
Date Rec'd 28/07/11
Date: 2/08/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B211080046	DIVERSEY TEGO 2000SC 50ml 25-42kGy	105	BATCH NO: W030188-CUST B/NO:372649 ACTUAL DOSE....35.4 - 36.4....kGy
B2FLE0010999 B211080046	SAMPLES	1	CUSTOMERS BATCH/CODE NO AS ORDER NO P017284 DATED 27.07.2011
Despatch Text: 2PLTS			
Total		106	Last Page 1 of 1
		All in accordance with current Technical Agreement	
		This is to certify that the above items have been irradiated as specified above	
		Authorising Signature: <i>J. Gifford</i>	
		For and on behalf of ISOTRON LIMITED	