



Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 24.09.11

Product Name	CLEARKLENS TEGO 2000SS VH25S
Product Code	7516427
Batch Number	10030673
Production Date	09/2011
Expiry Date	03/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Liquid	clear liquid
pH 100% (20°C)	JMD001	7.3 - 8.2	7.5
Specific Gravity (20°C)	JMD004	0.990 - 1.010	0.997

"SIGNATURE"

Gumirhe

Name *A. Gumirhe*

Function Name *Pharmaz Room Team Leader*

Please note Johnsondiversey did change the company name to Diversey and is in the phase of implementation

CERTIFICATE OF IRRADIATION

Number: **B**

0502811 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit B
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P017454
Date Rec'd 8/09/11
Date: 12/09/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010130 B211090315	TEGO SS STERILE FILL 25-40kGy	10	5 X B/No. W030670 - W030673 5 X B/No. W030671 ACTUAL DOSE...37.1 - 37.3...kGy CUSTOMER BATCH/CODE No.s AS ORDER No. P017454 DATED 7/9/11
	Despatch Text: 2PLTS		
	Total	10	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature: *J. Gaffney*

For and on behalf of ISOTRON LIMITED



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Winchester Road, Wickham, Fareham,
Hampshire PO17 5EU England
Telephone: +44(0)1329 832511
Fax: +44(0)1329 834262

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 30 Sep 11
Date Tested: 07 Oct 11
Date Completed: 21 Oct 11
Purchase Order: P017572

Test Article: Clearklens Tego 2000 SS VH25S
Sample Code: 2929
Batch Ref: WO30673
Qty Received: 20 x 900mL
Laboratory Reference: 0002477/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Product Standard Data Sheet	MM107/03	FM08
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 27 Oct 11 08:19:55

Mr K.A. Barker, CBiol, MSB
Business Manager - Microbiology

Consignment: 0002477
Lancaster

Certificate of Analysis - OSM

Print Number

P0000423

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited

ISO
17025

GMP GLP FDA
Good Manufacturing Practice Good Laboratory Practice Good Clinical Practice

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Winchester Road, Wickham, Fareham, Hampshire