



Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 24/9/11

Product Name	CLEARKLENS CLEANSINALD SS VH9S
Product Code	7516432
Batch Number	W030700
Production Date	09/2011
Expiry Date	03/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH (100% 20°C)	JDM001	9.0 - 11.7	10.7
Specific Gravity (20°C)	JDM004	0.990 - 1.010	0.997

"SIGNATURE"

Name

Caroline Percoe

Function Name

Pharmacy Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Winchester Road, Wickham, Fareham,
Hampshire PO17 5EU England
Telephone: +44(0)1329 832511
Fax: +44(0)1329 834262

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 28 Sep 11
Date Tested: 05 Oct 11
Date Completed: 19 Oct 11
Purchase Order: PO17555

Test Article: Cleansinald SS
Sample Code: 2981
Batch Ref: Work Order No. 30700
Qty Received: 20 x 900mL Bottles + 4 Spare
Laboratory Reference: 0002463/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Product Standard Data Sheet	MM107/03	FM06
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 24 Oct 11 15:00:02

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0002463
Lancaster

Certificate of Analysis - OSMM

Print Number

P0000392

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited