

Base last off with Filter



Diversey Europe BV

Maarssenbroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 11.02.2012

Product Name	CLEARLENs BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7514852
Batch Number	W031007
Production Date	02/2012
Expiration Date	02/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
pH (100% 20°C)	JDM001	11.5	13.5	11.9
Specific Gravity (20°C)	JDM004	1.015	1.035	1.017

On behalf of Diversey site Quality Manager	Name :	Matura
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7514852

Version : 03

Date of issuing : November 28th 2011

Base lost off with Filter



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 11.02.2012

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Product Code	7514852
Batch Number	W031007
Production Date	02/2012
Expiration Date	02/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Blue Liquid		Clear Blue Liquid
pH (100% 20°C)	JDM001	1.5	3.5	1.5
Specific Gravity (20°C)	JDM004	1.020	1.060	1.025

On behalf of Diversey site Quality Manager	Name :	Mature
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7514852	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

Remove for all with the
Base last off with filter



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 11.02.2012

Product Name	CLEARKLENS BI-SPORE COMBINED SOLUTION VH26
Product Code	7514852
Batch Number	W0310057
Production Date	02/2012
Expiration Date	02/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Yellow Liquid		Clear Yellow Liquid
pH (100% 20°C)	JDM001	2.5	5.5	2.5
Specific Gravity (20°C)	JDM004	1.000	1.010	1.001
Potential Chlorine Dioxide (ppm)		230	280	270ppm

On behalf of Diversey site Quality Manager	Name :	Matus
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7514852	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Synergy Health plc

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@synergyhealthplc.com
Website: www.synergyhealthplc.com

LA1 4XS

Cust. Ref: P018044
Date Rec'd 19/01/12
Date: 23/01/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010113 B212010477	JD-BOTTLES/CAPS/BAGS 25-40kGy	18	B/NO: W031006 X 4 W031007 . W031471 X 1 W031479 X 1 W031001 X 9 W031480 X 3 CUSTOMERS BATCH/CODE NO'S AS ORDER NO P018044 DATED 18.01.12
	Despatch Text: 6PLTS		
	Total	18	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature: *J. (offic)*

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 16 Feb 2012
Date Tested: 23 Feb 2012
Date Test Completed: 08 Mar 2012
Purchase Order: 18175

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0003619/5
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 16/02/2012
Test Article: Clearklens BI spore Base
Sample Code: Works Order 31007
Batch Ref: 315003
Qty Received: 20 x 100mL

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM05

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 14 Mar 2012 09:02:07

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003619

Certificate of Analysis - OSMM

Print Number: P0001997 Page 1 of 1



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 16 Feb 2012
Date Tested: 23 Feb 2012
Date Test Completed: 08 Mar 2012
Purchase Order: 18175

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0003619/6
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I; Ph Eur and USP
Date Received: 16/02/2012
Test Article: Clearklens BI Spore Activator
Sample Code: Works Order 31007
Batch Ref: 313004
Qty Received: 20 x 100mL

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM04

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 14 Mar 2012 09:02:16

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003619

Certificate of Analysis - OSMM

Print Number: P0001997

Page 1 of 1