

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 05.03.2012

Product Name	CLEARKLENS IPA VH1	
Product Code	7513400	
Batch Number	NO 314 85	
Production Date	02/2012	
Expiration Date	02/2014	- 102 - 110

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower -	Upper	
Appearance	Visual	Clear Colourless		clear colourless
		Lic	quid	Liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manager	Name :	Matura
	Position	Quality Control Inspector

This document being issued electronical	ly does not bear a signature	
COA 7513400	Version: 03	Date of issuing: November 28th 2011

CERTIFICATE OF IRRADIATION Number: B

0521995 00

FLE001

Flexible Medical Packaging Ltd

Unit 8

White Cross Industrial Estate

Hightown Lancaster

LA1 4XS

Cust. Ref: Date Rec'd P018220 23/02/12 27/02/12 Page

Synergy Health plc

Roydsdale Way · Euroway Trading Estate

Bradford · West Yorkshire BD4 6SE · United Kingdom Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061 Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

Date: 27/02/12			
ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B212020846	JD IPA BOTTLES 25-45kGy	354	B/No.W031485
			ROUTINE DOSE 35.5, 35.8, 39.1 kGy's
B2FLE0010999 B212020846	SAMPLES	5	
			CUSTOMER BATCH/CODE No.s AS ORDER No.P018220 DATED 22/2/12
	Despatch Text:		
	7PLTS		
	Total	359	Last Page 1 of 1
		All in acco	ordance with current

Technical Agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories Limited, Winchester Road, Wickham, Fareham, Hampshire P017 5EU England Telephone: +44(0)1329 832511

Fax:

+44(0)1329 834262

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8

Hightown

White Cross Industrial Estate

Lancaster

LA14XS

Date Received:

05 Mar 2012

Date Tested:

07 Mar 2012

Date Test Completed:

21 Mar 2012

Purchase Order:

18274

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0003806/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received: Test Article: 05/03/2012

Sample Code:

Clearkiens IPA Block 1

Batch Ref:

3200 WO31485

Qty Received:

20 x 900mLs + Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Elmyn

Approval is provided by Electronic Signature. Their name is shown below.

Date: 2

Date: 22 Mar 2012 15:02:07

Mr K.A. Barker, CBiol, MSB

Business Manager - Microbiology

Consignment: 0003806

Certificate of Analysis - OSMM

Print Number; P0002147

Page 1 of 1







Wickham Laboratories Limited, Winchester Road, Wickham, Fareham, Hampshire P017 5EU England Telephone: +44(0)1329 832511

ax: +44(0)1329 834262

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Mr J Jackson

Flexible Medical Packaging Limited

Unit 8

Hightown

White Cross Industrial Estate

Lancaster

LA1 4XS

Date Received:

05 Mar 2012

Date Tested:

06 Mar 2012

Date Test Completed:

06 Mar 2012

Purchase Order:

18274

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0003806/2

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received:

05/03/2012

Test Article:

Clearklens IPA Block 1

Sample Code:

3200

Batch Ref: Qtv Received: Works No. 31485

1 x 900mL

A STATE OF THE PROPERTY OF THE			
Test	Method Item	Result	
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL	
Spiked Recovery	MM110/01	84 %	
Dilution Tested	MM110/04	1/40	
Endotoxin Spike Level	MM110/02	0.5 EU/mL	
Product Standard Data Sheet	FG047/psd	FM02	

Comments

Test carried out according to current Pharmacopoelas.

E. Gorya.

Approval is provided by Electronic Signature. Their name is shown below,

Date: 07 Mar 2012 11:23:23

Mrs P. Pham-Lengoc, BSc, MSB

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0003806

Certificate of Analysis - OSMM

Print Number: P0001919

Page 1 of 1