



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 27.04.13

Product Name	CLEARLENS DE VH29
Product Code	6069643
Batch Number	FMP13 095 33893
Production Date	04/2013
Expiration Date	04/2015

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear Colourless liquid
Specific Gravity (20°C)	JDM004	0.880	0.888	0.885

On behalf of Diversey site Quality Manager	Name :	A. Matena, C. Bae
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 6069643	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 23-Apr-2013

UK32S11014023-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

ISO 11137:2006 Sterilisation of Health Care Products
 ISO 9001:2008 Quality Management System
 ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
 Unit 8, White Cross Ind Estate
 Hightown
 Lancaster
 Lancashire LA1 4XS
 UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Customer Reference Number:	P020231
Product Description:	JD DE 900ML BOTTLES 25-45kGy
Validation Reference:	4.0624
Quantity Received:	708
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	702 X B/No.FMP13095/33893 + 6 SAMPLES , 13 PLTS LOT 1 THURS 11.4.13
Shop Order Number:	S11014023-1-1
Synergy Health Sales Part Reference:	1002733

Irradiation Data

Date and Time of Irradiation:	23-Apr-2013 05:04
Reference Dose Range kGy:	36.7 - 38.2
Calculated Minimum Dose kGy:	26.1
Calculated Maximum Dose kGy:	40.4

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories
contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 29 Apr 2013
Date Tested: 02 May 2013
Date Test Completed: 16 May 2013
Purchase Order: P020336

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0010019/2
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 29/04/2013
Test Article: ClearKlens DE Block 1
Sample Code: FMP13095 33893
Batch Ref: 4347
Qty Received: 20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 21 May 2013 14:24:26

Mr K.A. Barker, CBiol, MSB
Business Manager - Microbiology

Consignment: 0010019

Certificate of Analysis - OSMM

Print Number: P0010647

Page 1 of 1



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 29 Apr 2013
Date Tested: 02 May 2013
Date Test Completed: 02 May 2013
Purchase Order: PO20336

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0010019/1
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 29/04/2013
Test Article: ClearKlens DE Block 1
Sample Code: -
Batch Ref: FMP13095 33893
Qty Received: 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL
Spiked Recovery	MM110/01	96 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM01

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 03 May 2013 15:21:51

Mr K.A. Barker, CBiol, MSB
Business Manager - Microbiology

Consignment: 0010019

Certificate of Analysis - OSMM

Print Number: P0010304

Page 1 of 1