

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 11/7/13

Product Name	CLEARKLENS IPA VH1	
Product Code	7514372	
Batch Number	FMP13176, 34453	
Production Date	06/2013	
Expiration Date	06/2015	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless	year colonies
••		Liquid	clear colonies
Specific Gravity (20°C)	JDM004	0.872 0.883	0.477
		* .	

On behalf of Diversey site Quality Manager	Name:	Char	, A Marken
	Position	Quality Contro	ol Inspector

This document being issued electronically does not bear a signature				
COA 7514372	Version: 03	Date of issuing: November 28th 2011		



Wickham Laboratories Limited, Hoeford Point, Barwell Lane, Gosport Hampshire P013 0AU England

Telephone: +44(0)1329 226600

+44(0)1329 832511

Fax:

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8

Hightown

White Cross Industrial Estate

Lancaster

LA1 4XS

Date Received:

11 Jul 2013

Date Tested:

16 Jul 2013

Date Test Completed:

30 Jul 2013

Purchase Order:

20680

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0011252/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

11/07/0

Date Received:

11/07/2013

Clearklens IPA Block 2

Test Article: Sample Code:

Batch Ref:

FMP13176 34453

Qty Received:

20 x 900mL bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

MY3~

Date: 01 Aug 2013 08:48:32

Mr K.A. Barker, CBiol, MSB

Business Manager - Microbiology

Consignment: 0011252

Certificate of Analysis - OSMM

Print Number: P0012404

Page 1 of 1





http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 05-Jul-2013

UK32S11032613-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

ISO 11137:2006 Sterilisation of Health Care Products

ISO 9001:2008 Quality Management System

ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS **UNITED KINGDOM**

All in accordance with current Technical Agreement

Order Information		
Customer Reference Number:	P020635	
Product Description:	JD IPA BOTTLES 25-45kGy	
Validation Reference:	4.0621	
uantity Received: 298		
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy: 45.0 *		
Customer Unit Lot/Batch Number:	293 X B/No.FMP13176/34453 + 5 SAMPLES , 6 PLTS LOT 2 MON 1.7.13	
Shop Order Number:	S11032613-1-1	
Synergy Health Sales Part Reference:	1002734	
Irra	adiation Data	
Date and Time of Irradiation:	05-Jul-2013 16:54	
Reference Dose Range kGy:	36.8 - 37.0	
Calculated Minimum Dose kGy:	26.6	
Calculated Maximum Dose kGy:	39.3	

*FMP13176 34453 # Gaugn 17.07.13



Wickham Laboratories Limited, Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600

+44(0)1329 832511

Fax:

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8

Hightown White Cross Industrial Estate

Lancaster

LA1 4XS

Date Received:

11 Jul 2013

Date Tested:

15 Jul 2013

Date Test Completed:

15 Jul 2013

Purchase Order:

20680

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0011252/2

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received: Test Article:

11/07/2013

Clearklens IPA Block 2

Sample Code:

Batch Ref:

FMP13176 34453

1 x 900mL bottle Qty Received:

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.20 EU/mL
Spiked Recovery	MM110/01	107 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02
Name and the second of the sec		

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Sher Wood

Date: 17 Jul 2013 15:31:48

Mrs S.L. Wood, BSc (Hons), CBiol, MSB

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0011252

Certificate of Analysis - OSMM

Print Number: P0012013

Page 1 of 1