



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 30/9/13

Product Name	CLEARKLENS DE VH29
Product Code	7515783
Batch Number	FMP13254 34575
Production Date	09/2013
Expiration Date	09/2015

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless liquid
Specific Gravity (20°C)	JDM004	0.880	0.888	0.885

On behalf of Diversey site Quality Manager	Name :	C. Piscoe, A. Mouton
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7515783	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 25-Sep-2013

UK32S11054923-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
 Unit 8, White Cross Ind Estate
 Hightown
 Lancaster
 Lancashire LA1 4XS
 UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002733
Customer Reference Number:	P021068
Product Description:	JD DE 900ML BOTTLES 25-45kGy
Validation Reference:	4.0624
Quantity Received:	704
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	699 X B/No.FMP13254 34575 + 5 SAMPLES , 13 PLTS LOT 1 THURS 19.9.13

Irradiation Data

Date and Time of Irradiation:	24-Sep-2013 20:11
Reference Dose Range kGy:	35.5 - 37.1
Calculated Minimum Dose kGy:	25.3
Calculated Maximum Dose kGy:	39.2

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069

Ms D Henderson
 Flexible Medical Packaging Limited
 Unit 8
 Hightown
 White Cross Industrial Estate
 Lancaster
 LA1 4XS

Date Received: 30 Sep 2013
Date Tested: 03 Oct 2013
Date Test Completed: 03 Oct 2013
Purchase Order: 21096

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0012571/1
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 30/09/2013
Test Article: Clearklens DE Block 2
Sample Code: -
Batch Ref: FMP 13254 34575
Qty Received: 1 Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL
Spiked Recovery	MM110/01	165 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM01

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.



Date: 04 Oct 2013 16:15:19

 Mrs P. Pham-Lengoc
 Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0012571

Certificate of Analysis - OSMM

Print Number: P0013935

Page 1 of 1

 ISO
17025

GLP GMP FDA

 Company Registered in England No. 752951 Wickham Laboratories Limited,
 Registered Office: Winchester Road, Wickham, Fareham, Hampshire



Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 30 Sep 2013
Date Tested: 05 Oct 2013
Date Test Completed: 19 Oct 2013
Purchase Order: 21096

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0012571/2
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 30/09/2013
Test Article: Clearklens DE Block 2
Sample Code: -
Batch Ref: FMP 13254 34575
Qty Received: 20 x 900mL bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

See Non-Conformance Report PHM 2013 249
Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 21 Oct 2013 16:26:20

Mrs S.L. Wood
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0012571

Certificate of Analysis - OSMM

Print Number: P0014315

Page 1 of 1

ISO
17025

GLP GMP FDA

Company Registered in England No. 752951 Wickham Laboratories Limited,
Registered Office: Winchester Road, Wickham, Fareham, Hampshire