

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: 20.09.2013

Product Name	CLEARKLENS TEGO 2000SC VH25 7513387		
Product Code			
Batch Number	FMP 122/17, 3/1205		
Production Date	09 1202		
Expiration Date	00 13015		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test Method		The state of the s	Results
Visual	Clear Slig		Chech slightly tellow Liquid
JDM001	7.5		7.6
JDM004	0.993	1.003	1.001
	Visual JDM001	Visual Clear Sligh Liq JDM001 7.5	Visual Clear Slightly Yellow Liquid  JDM001 7.5 8.5

On behalf of Diversey site  Quality Manager	Name:	A. Matera, O. Kinchner
	Position	Quality Control Inspector

This document being issued	electronically does not bear a signature	
COA 7513387	Version: 03	<u> </u>
0011 / 31330/	version: 03	Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

## **Certificate of Irradiation**

Date Issued: 16-Sep-2013

UK32S11052713-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information				
Account Number: Synergy Health Sales Part Reference: Customer Reference Number: Product Description: Validation Reference: Quantity Received: Customer Minimum Specification kGy: Customer Maximum Specification kGy: Customer Unit Lot/Batch Number:	100432 1002722 P021019 DIVERSEY TEGO 2000SC 50ml 25-42kGy 4.0716 107 25.0 42.0 105 X FMP13247 34805 + 2 SAMPLES, 2 PLTS LOT 1 THURS 12.9.13			
Irradiation Data				
Date and Time of Irradiation: Reference Dose Range kGy: Calculated Minimum Dose kGy: Calculated Maximum Dose kGy:	14-Sep-2013 21:00 33.3 - 36.4 27.7 40.0			



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: Alt. Telephone: +44(0)1329 226600 +44(0)1329 832511

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

23 Sep 2013

**Date Tested:** 

26 Sep 2013

**Date Test Completed:** 

21 Oct 2013

**Purchase Order:** 

P021084

## **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 

0012459/1

**Test Required:** 

Sterility Test by Membrane Filtration (Steritest) Method I with Subculture:

PH Eur and USP

**Date Received:** 

23/09/2013

Test Article: Sample Code: TEGO 2000SC Block 1

FMP13247 34805

Batch Ref:

435211

**Qty Received:** 20 x 50mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Growth of Sub-culture in Tryptone Soya Broth at 30-35°C after 4 days	MM107/07	No growth in two broths
Growth of Sub-culture in Fluid Thioglycollate at 20-25°C after 4 days	MM107/08	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM08

## **Comments**

See Non-Conformance Report PHM 2013 270

Approval is provided by Electronic Signature. Their name and position is shown below.

Ens Loos

Date: 23 Oct 2013 08:47:22

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0012459

Certificate of Analysis - OSMM

Print Number: P0014340

Page 1 of 1

ISO 17025 GLP GMP FDA

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Winchester Road, Wickham, Fareham, Hampshire