

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 27, 11, 13

Product Name	CLEARKLENS IPA VH1	
Product Code	7513400	
Batch Number	FMP 13311 35052	
Production Date	11/2013	
Expiration Date	11/2015	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lower -	mits Upper	Results
Appearance	Visual	4	olourless quid	clear Colourless
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875

O-1-1-16-6D:	Name:	Matero, Patyighe
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector

This document	being	issued	ele	ctroni	cally	v de	oes	not	bear	a sig	mature	

COA 7513400 Version: 03 Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 22-Nov-2013

UK32S11074532-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information Account Number: 100432 Synergy Health Sales Part Reference: 1002734 **Customer Reference Number:** P021336 **Product Description:** JD IPA BOTTLES 25-45kGy Validation Reference: 4.0621 **Quantity Received:** 706 25.0 Customer Minimum Specification kGy: Customer Maximum Specification kGy: 45.0 701 x B/N FMP13311 35052 + 5 SAMPLES; 13 Customer Unit Lot/Batch Number: PLTS, LOT 1, THUR 14.11.13 Irradiation Data Date and Time of Irradiation: 22-Nov-2013 05:31 37.6 - 39.3 Reference Dose Range kGy: Calculated Minimum Dose kGy: 27.2 41.7 Calculated Maximum Dose kGy:

Irradiation Release Authorised By Synergy Health plc



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

28 Nov 2013

Date Tested:

03 Dec 2013

Date Test Completed:

17 Dec 2013

Purchase Order:

21395

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0013604/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received:

28/11/2013

Test Article:

CLEARKLENS IPA BLOCK 1

Sample Code:

FMP13311 35052

Batch Ref:

4951

Qty Received:

20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

amore

Date: 18 Dec 2013 18:13:01

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0013604

Certificate of Analysis - OSMM

Print Number: P0016213

Page 1 of 1

ISO 17025



Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Winchester Road, Wickham, Fareham, Hampshire



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone:

+44(0)1329 226600 Alt. Telephone: +44(0)1329 832511

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

28 Nov 2013

Date Tested:

02 Dec 2013

Date Test Completed:

02 Dec 2013

Purchase Order:

21395

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0013604/2

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received:

28/11/2013

Test Article:

CLEARKLENS IPA BLOCK 1

Sample Code:

4951

Batch Ref: Qty Received: FMP13311 35052 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	122 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Ens Wood

Date: 04 Dec 2013 21:46:14

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0013604

Certificate of Analysis - OSMM

Print Number: P0015702

Page 1 of 1

