Diversey

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 03.01.2014

Product Name	CLEARKLENS IPA VH1	
Product Code	7514372	
Batch Number	FMP13343 35386	
Production Date	12/2013	
Expiration Date	12/2015	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin	nits	Results
		Lower -	Upper	
Appearance	Visual	Clear C	olourless	clear colorvess
		Lic	quid	liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875

	Name:	Anakere, Parlysisto
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector

This document being issued electronical	ly does not bear a signature	
COA 7514372	Version: 03	Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 23-Dec-2013

UK32S11089932-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information Account Number: 100432 Synergy Health Sales Part Reference: 1002734 **Customer Reference Number:** P021513 **Product Description:** JD IPA BOTTLES 25-45kGy Validation Reference: 4.0621 **Quantity Received:** 707 Customer Minimum Specification kGy: 25.0 Customer Maximum Specification kGy: 45.0 Customer Unit Lot/Batch Number: LOT 4; THUR 19.12.13; 13 PLTS; B/NO. FMP13343 35386 x 702 + 5 SAMPLES-RE ISSUE 23.12.13

Irradiation Data		
Date and Time of Irradiation:	20-Dec-2013 19:33	
Reference Dose Range kGy:	36.3 - 38.3	
Calculated Minimum Dose kGy:	26.2	
Calculated Maximum Dose kGy:	40.6	



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: Alt. Telephon +44(0)1329 226600 +44(0)1329 832511

Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

06 Jan 2014

Date Tested:

15 Jan 2014

Date Test Completed:

29 Jan 2014

Purchase Order:

PO21550

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0014143/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received:

06/01/2014

Test Article:

Clearklens IPA Block 2

Sample Code:

FMP13343 35386

Batch Ref:

5047

Qty Received:

20 x 900mL bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

1 Bottle found leaking on receipt; 1 replacement bottle received 10 Jan 2014.

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Ens Loss.

Date: 30 Jan 2014 15:45:32

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0014143

Certificate of Analysis - OSMM

Print Number: P0017148

Page 1 of 1

ISO 17025



Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Winchester Road, Wickham, Fareham, Hampshire



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: Alt. Telephone: +44(0)1329 226600 +44(0)1329 832511

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

06 Jan 2014

Date Tested:

09 Jan 2014

Date Test Completed:

09 Jan 2014

Purchase Order:

PO21550

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0014143/2

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received:

06/01/2014

Test Article:

Clearklens IPA Block 2

Sample Code: Batch Ref: FMP13343 35386 5047

Qty Received:

1 x 900mL bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	85 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Sun Lass.

Date: 10 Jan 2014 14:16:48

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0014143

Certificate of Analysis - OSMM

Print Number: P0016686

Page 1 of

ISO 17025

