

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 20/12/13

Product Name	CLEARKLENS TEGO 2001SC VH24
Product Code	7517177
Batch Number	FMP13336 35504
Production Date	12/12/013
Expiration Date	12/2015

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test Method	Lin Lower -		Results
Visual		htly Yellow	Clear Slightly
JDM001	8.5	9.5	yellow 1, good
JDM004	1.040	1.080	1.054
	Visual JDM001	Visual Clear Sligi Liq JDM001 8.5	Visual Clear Slightly Yellow Liquid JDM001 8.5 9.5

On behalf of Diversey site Quality Manager	Name:	Place, Olpa kinchma.
	Position	Quality Control Inspector



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 16-Dec-2013

UK32S11086856-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information			
Account Number: Synergy Health Sales Part Reference: Customer Reference Number: Product Description: Validation Reference: Quantity Received: Customer Minimum Specification kGy: Customer Maximum Specification kGy: Customer Unit Lot/Batch Number:	100432 1002723 P021479 DIVERSEY TEGO2001 10x5x100 25-42kGy 4.0706 106 25.0 42.0 LOT 2; 2 PLTS; THUR 12.12.13; 104 x B/NO. FMP 13336-35504 + 2 SAMPLES		
	rradiation Data		
Date and Time of Irradiation: Reference Dose Range kGy: Calculated Minimum Dose kGy: Calculated Maximum Dose kGy:	15-Dec-2013 01:18 31.8 - 32.5 26.2 35.2		