Diversey Europe Operations BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427



CERTIFICATE OF ANALYSIS

Date: 7/3/14

Product Name	CLEARKLENS TEGO 2000SS VH25S	
Product Code	7516428	
Batch Number	FMP14048 35724	
Production Date	02/2014	
Expiration Date	08/2015	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin Lower -	nits Upper	Results
Appearance	Visual		htly Yellow Juid	gellow lighty
pH 100% (20°C)	JDM001	6.2	8.2	7.4
Specific Gravity (20°C)	JDM004	0.990	1.010	0.994

On behalf of Diversey site Quality Manager	Name :	Masor Party Sto
	Position	Quality Control Inspector

1 nis	document	being issued	l electronical	ly does	not	bear	a signature	
				-				

COA VH25S	Version: 05	Date of issuing: June 11th 2013
-----------	-------------	---------------------------------



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 28-Feb-2014

UK32S11116421-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information			
Account Number:	100432		
Synergy Health Sales Part Reference:	1035147		
Customer Reference Number:	P021806		
Product Description:	TEGO 2000SS BLK1 25-45KGY		
Validation Reference:	4.0621		
Quantity Received:	26		
Customer Minimum Specification kGy:	25.0		
Customer Maximum Specification kGy:	45.0		
Customer Unit Lot/Batch Number:	LOT 2; 1 PLT; THUR 20.02.14; 21 x B/NO.		
	FMP14048 35724 + 5 SAMPLES		
	rradiation Data		
Date and Time of Irradiation:	28-Feb-2014 06:00		
Reference Dose Range kGy:	38.9 - 40.6		
Calculated Minimum Dose kGy:	28.0		
Calculated Maximum Dose kGy:	43.0		



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

10 Mar 2014

Date Tested:

12 Mar 2014

Date Test Completed:

03 Apr 2014

Purchase Order:

PO21894

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0015392/2

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I with Subculture:

PH Eur and USP

Date Received:

10/03/2014

Test Article: Sample Code: TEGO 2000SS Block 2

FMP14048 35724 5233

Batch Ref: Qty Received:

21 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Growth of Sub-culture in Tryptone Soya Broth at 30-35°C after 4 days	MM107/07	No growth in two broths
Growth of Sub-culture in Fluid Thioglycollate at 20-25°C after 4 days	MM107/08	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM08

Comments

21 x 900mL bottles received As Steritest Sterility Test

Approval is provided by Electronic Signature. Their name and position is shown below.

amore

Date: 04 Apr 2014 16:26:15

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0015392

Certificate of Analysis - OSMM

Print Number: P0018872

Page 1 of 1

ISO GLP GMP FDA

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Winchester Road, Wickham, Fareham, Hampshire