

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 21.05.14

Product Name	CLEARKLENS CLEANSINALD SS VH9S		
Product Code	7516431		
Batch Number	FMP 14 141 36015		
Production Date	05/2014		
Expiration Date			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin Lower -	nits Upper	Results
Appearance	Visual	-	htly Yellow	clear slightly,
-II (4000) 2000)		Liq	uid	Yellow Liquid
pH (100% 20°C)	JDM001	9.0	11.4	10.6
Specific Gravity (20°C)	JDM004	0.990	1.010	0.998

On behalf of Diversey site Quality Manager	Name:	Masera, E Garyon.
	Position	Quality Control Inspector

Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 25-Mar-2014

UK32S11127708-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002720
Customer Reference Number:	P021912
Product Description:	C/SINALD SS 900ml BOTTLE 25-40kGy
Validation Reference:	4.0615
Quantity Received:	5
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	40.0
Customer Unit Lot/Batch Number:	LOT 7; 1 PLT; THUR 13.03.14; B/NO. W036016

Irradiation Data

Date and Time of Irradiation:	25-Mar-2014 22:30
Reference Dose Range kGy:	34.8 - 34.9
Calculated Minimum Dose kGy:	28.6
Calculated Maximum Dose kGy:	36.1

Items irradiated under WO36016 will be used in finished product batch code

FMP14141 36015 É. Craujus 22-05-14



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511

Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

22 May 2014

Date Tested:

24 May 2014

Date Test Completed:

07 Jun 2014

Purchase Order:

22316

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0016731/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received:

22/05/2014

Test Article:

Clearklens Cleansinald SS

Sample Code:

5376

Batch Ref:

FMP14141 36015

Qty Received:

20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM06

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 09 Jun 2014 08:03:45

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0016731

GLP GMP FDA

Certificate of Analysis - OSMM

Print Number: P0020447

Page 1 of 1

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Winchester Road, Wickham, Fareham, Hampshire