

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 28/7/14

Product Name	CLEARKLENS DE VH29	
Product Code	6069643	
Batch Number	FMP14185 36363	
Production Date	07/2014	
Expiration Date	07/2016	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower -	Upper	
Appearance	Visual	Clear Colourless		lieur colourles
		Lic	quid	11201d
Specific Gravity (20°C)	JDM004	0.880	0.888	0.883
			-	

0.1.1.16.60	Name:	Classe Bows
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector

This document being issued electronical	ly does not bear a signature	
COA 6069643	Version: 03	Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 21-Jul-2014 UK32S11198008-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information **Account Number:** 100432 **Synergy Health Sales Part Reference:** 1002733 **Customer Reference Number:** P022562 **Product Description:** JD DE 900ML BOTTLES 25-45kGy Validation Reference: 4.0624 703 **Quantity Received: Customer Minimum Specification kGy:** 25.0 **Customer Maximum Specification kGy:** 45.0 **Customer Unit Lot/Batch Number:** 699 X B/NO.FMP14185 36363 + 4 SAMPLES THUR 10.07.14 13 PLTS LOT 2 Other Process Details: Released against validation 4.1756. Calculated minimum dose = 26.1kGy. Calculated maximum dose = 43.0kGyIrradiation Data Date and Time of Irradiation: 18-Jul-2014 07:09



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

28 Jul 2014

Date Tested:

30 Jul 2014

Date Test Completed:

13 Aug 2014

Purchase Order:

PO22639

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0017832/1

Test Required: Date Received: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

28/07/2014

Test Article:

CLEARKLENS DE BLOCK 1

Sample Code:

Batch Ref: Qty Received: FMP14185 36363 20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 14 Aug 2014 14:11:23

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0017832

GLP GMP FDA

Certificate of Analysis - OSMM

Print Number: P0021952

Page 1 of 1

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone:

+44(0)1329 226600 Alt. Telephone: +44(0)1329 832511

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

28 Jul 2014

Date Tested:

01 Aug 2014

Date Test Completed:

01 Aug 2014

Purchase Order:

PO22639

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0017832/2

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received:

28/07/2014

Test Article:

CLEARKLENS DE BLOCK 1

Sample Code:

Batch Ref:

FMP14185 36363

Qty Received:

1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL
Spiked Recovery	MM110/01	85 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM01

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 04 Aug 2014 10:39:48

Mr K.A. Barker

Business Manager - Microbiology

Consignment: 0017832

Certificate of Analysis - OSMM

Print Number: P0021748

Page 1 of 1

17025 GLP GMP FDA

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire