



Diversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 25/11/14

Product Name	CLEARKLENS IPA VH1
Product Code	7513400
Batch Number	EMP14307 37087
Production Date	11/2014
Expiration Date	11/2016

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875

On behalf of Diversey site Quality Manager	Name :	Chloe O. Linchew
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513400	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 17-Nov-2014

UK32S11277269-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002734
Customer Reference Number:	P023283
Product Description:	JD IPA BOTTLES 25-45kGy
Validation Reference:	4.1723
Quantity Received:	708
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	704 X B/N FMP14307 37087 + 4 SAMPLES THUR 13.11.14 13 PLTS LOT 4

Irradiation Data

Date and Time of Irradiation:	16-Nov-2014 03:34
Reference Dose Range kGy:	36.8 - 37.4
Calculated Minimum Dose kGy:	26.0
Calculated Maximum Dose kGy:	40.0

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Alt. Telephone: +44(0)1329 832511
Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 24 Nov 2014
Date Tested: 25 Nov 2014
Date Test Completed: 25 Nov 2014
Purchase Order: 23348

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0019933/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 24/11/2014
Test Article: ClearKlens IPA Block 1
Sample Code: 6022
Batch Ref: FMP14307 37087
Qty Received: 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	84 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

C Moore

Date: 26 Nov 2014 15:58:35

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0019933

Certificate of Analysis - OSMM

Print Number: P0024041

Page 1 of 1

ISO
17025

GLP GMP FDA

Company Registered in England No. 752951 Wickham Laboratories Limited,
Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Alt. Telephone: +44(0)1329 832511
Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 24 Nov 2014
Date Tested: 28 Nov 2014
Date Test Completed: 12 Dec 2014
Purchase Order: 23348

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0019933/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 24/11/2014
Test Article: ClearKlens IPA Block 1
Sample Code: 6022
Batch Ref: FMP14307 37087
Qty Received: 20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

C Moore

Date: 15 Dec 2014 16:34:07

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0019933

Certificate of Analysis - OSMM

Print Number: P0024440

Page 1 of 1

ISO
17025

GLP GMP FDA

Company Registered in England No. 752951 Wickham Laboratories Limited,
Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire