

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 27/11/14

Product Name	CLEARKLENS CLEANSINALD SC VH9		
Product Code	7516429		
Batch Number	FMP14331 37487		
Production Date	11/2014		
Expiration Date	11/2016		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lower -	nits Upper	Results
Appearance	Visual	1	htly Yellow	yellow lightly
pH (100% 20°C)	JDM001	12.0	13.0	13.0
Specific Gravity (20°C)	JDM004	1.040	1.060	1.050

On behalf of Diversey site Quality Manager	Name:	Associ Ospa Lindina
	Position	Quality Control Inspector

This document being issued e	lectronically does not bear a signature	
COA 7516429	Version: 03	Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 02-Nov-2014

UK32S11268602-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS **UNITED KINGDOM**

All in accordance with current Technical Agreement

Account Number: 100432 Synergy Health Sales Part Reference: 1002719 **Customer Reference Number:** P023201 C/SINALD BLK BOTT/CAPS/BAGS 25-40kGy **Product Description:** Validation Reference: 4.1666 **Quantity Received:** 11 Customer Minimum Specification kGy: 25.0 Customer Maximum Specification kGy: 40.0 Customer Unit Lot/Batch Number: B/N W037192 THUR 30.10.14 2 PLTS LOT 3

Order Information

Irradiation Data		
Date and Time of Irradiation:	02-Nov-2014 20:59	
Reference Dose Range kGy:	33.1 - 33.5	
Calculated Minimum Dose kGy:	26.7	
Calculated Maximum Dose kGy:	34.4	

Items irradiated under WO37192 will be used in finished product batch code FMP14331 37487 £ Garage 11.12.14



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone:

+44(0)1329 226600 Alt. Telephone: +44(0)1329 832511

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

08 Dec 2014

Date Tested:

10 Dec 2014

Date Test Completed:

24 Dec 2014

Purchase Order:

23417

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0020180/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received: Test Article:

08/12/2014 Cleansinald SC

Sample Code:

483397

Batch Ref: Qty Received: FMP14331 37487

20 x 25mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested Product Standard Data Sheet	MM107/05 FG047/psd	20 x 25 mL FM06

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 29 Dec 2014 14:50:33

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0020180

Certificate of Analysis - OSMM

Print Number: P0024705

Page 1 of 1

GLP GMP FDA

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire