

**Diversey Europe BV** 

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

## **CERTIFICATE OF ANALYSIS**

Date: 6.03.2015

Product Name	CLEARKLENS TEGO 2001SS VH24S		
Product Code	7513403		
Batch Number	FMP 15 040 37604		
Production Date	02/2015		
Expiration Date	12/2015 -		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	<b>Test Method</b>	Limits		Results
		Lower -	Upper	
Appearance	Visual			clear Slightly
		Liquid		Yellow Liquid
pH 100% (20°C)	JDM001	8.0	9.0	8.7
Specific Gravity (20°C)	JDM004	0.980	1.020	0.994

On behalf of Diversey site Quality Manager	Name :	Palyish Maken	ے
	Position	Quality Control Inspector	

This document being issued electronically does not bear a signature		
COA 7513403	Version: 04	Date of issuing: April 19th 2012



http://www.synergyhealthplc.com

# **Certificate of Irradiation**

Date Issued: 16-Feb-2015 UK32S11329987-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

Date and Time of Irradiation:

All in accordance with current Technical Agreement

#### Order Information **Account Number:** 100432 Synergy Health Sales Part Reference: 1002758 **Customer Reference Number:** P023778 **Product Description:** TEGO 2001 SS 25-45kGy Validation Reference: 4.0968 **Quantity Received:** 120 Customer Minimum Specification kGy: 25.0 Customer Maximum Specification kGy: 45.0 Customer Unit Lot/Batch Number: 115 X B/NO: FMP15040 37604 THURS 12 FEB 2015 LOT 3 2PLTS + 5 SAMPLES Other Process Details: Released against validation 4.1903. Calculated minimum dose 26.4 kGy and calculated maximum dose 40.3 kGy **Irradiation Data**

15-Feb-2015 07:46

Irradiation Release Authorised By Synergy Health plc



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

24 Feb 2015

Date Tested:

26 Feb 2015

Date Test Completed:

20 Mar 2015

Purchase Order:

23860

## **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 

0021335/1

**Test Required:** 

Sterility Test by Membrane Filtration (Steritest) Method I with Subculture:

PH Eur and USP

**Date Received:** 

24/02/2015

Test Article:

**TEGO 2001ss** 

Sample Code:

**Batch Ref:** 

FMP15040 37604

**Qty Received:** 21 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Growth of Sub-culture in TSB at 20-25°C after 4 - 7 days	MM107/07	No growth in one broth
Growth of Sub-culture in THY at 30-35°C after 4 - 7 days	MM107/08	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM09

### Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Sur Load

Date: 23 Mar 2015 17:06:14

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0021335

Certificate of Analysis - OSMM

Print Number: P0026268

Page 1 of 1

