

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 9.04.2015

CLEARKLENS DE VH29		
7515783		
FMP 15047 37691		
23/02/2015		
23 02 2017		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits Lower - Upper	Results
Appearance	Visual		Clear Colorle
Specific Gravity (20°C)	JDM004	0.880 0.888	0.883

On behalf of Diversey site	Name:	Harry igla Makero
Quality Manager	Position	Quality Control Inspector

This document being issued electronical	ly does not bear a signature	
COA 7515783	Version: 03	Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 11-Mar-2015

UK32S11339023-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

Calculated Maximum Dose kGy:

All in accordance with current Technical Agreement

Order Information **Account Number:** 100432 Synergy Health Sales Part Reference: 1002733 **Customer Reference Number:** P023887 **Product Description:** JD DE 900ML BOTTLES 25-45kGy Validation Reference: 4.1756 **Quantity Received:** 703 Customer Minimum Specification kGy: 25.0 Customer Maximum Specification kGy: 45.0 Customer Unit Lot/Batch Number: 699 X B/N FMP15047 37691 + 4 SAMPLES THUR 26 FEB 2015 13 PLTS LOT 1 **Irradiation Data** 10-Mar-2015 14:26 Date and Time of Irradiation: Reference Dose Range kGy: 37.5 - 38.4 Calculated Minimum Dose kGy: 28.1

41.6

Irradiation Release Authorised By Synergy Health plc



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: Alt. Telephone:

+44(0)1329 226600 +44(0)1329 832511

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

16 Mar 2015

Date Tested:

19 Mar 2015

Date Test Completed:

02 Apr 2015

Purchase Order:

24004

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0021676/2

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received:

16/03/2015

Test Article:

ClearKlens DE Block 2

Sample Code:

6285

Batch Ref:

FMP15047 37691 20 x 900mL Bottles

Qty Received:

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested Product Standard Data Sheet	MM107/05 FG047/psd	20 x 50 mL FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 07 Apr 2015 15:33:35

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0021676

Certificate of Analysis - OSMM

Print Number: P0026510





Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received: 16 Mar 2015

Date Tested: 17 Mar 2015

Date Test Completed: 17 Mar 2015

Purchase Order: 24004

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0021676/1

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Date Received:

16/03/2015

Test Article:

ClearKlens DE Block 2

Sample Code:

6285

Batch Ref: Qty Received: FMP15047 37691 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL
Spiked Recovery	MM110/01	146 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM01

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Missen

Date: 19 Mar 2015 12:56:22

Mr K.A. Barker

Business Manager - Microbiology

Consignment: 0021676

GLP GMP FDA

Certificate of Analysis - OSMM

Print Number: P0026208

Page 1 of 1

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire