



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 10.05.2016

Product Name	CLEARKLENS TEGO 2000SC VH25
Product Code	7513387
Batch Number	FHP 16111 - 40355
Production Date	20/04/2016
Expiration Date	EXP 20/04/2019

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid		Clear Slightly Yellow Liquid
pH 100% (20°C)	JDM001	7.5	8.5	4.5
Specific Gravity (20°C)	JDM004	0.993	1.003	0.999

On behalf of Diversey site Quality Manager	Name :	H. Harabach Paulysch
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513387	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 07-May-2016

UK32S11613791-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002722
Customer Reference Number:	PO26467
Product Description:	DIVERSEY TEGO 2000SC 50ml 25-42kGy
Validation Reference:	4.1970
Quantity Received:	107
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	42.0
Customer Unit Lot/Batch Number:	106 X FMP BATCH NO: FMP16111 40355 + 1 SAMPLE THURS 28TH APR 2016 2 PLTS LOT 3

Irradiation Data

Date and Time of Irradiation:	07-May-2016 00:09
Reference Dose Range kGy:	33.9 - 34.0
Calculated Minimum Dose kGy:	27.5
Calculated Maximum Dose kGy:	37.5

Irradiation Release Authorised By Synergy Health plc

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 11 May 2016
Date Tested: 17 May 2016
Date Test Completed: 08 Jun 2016
Purchase Order: 26553

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0028900/1
Test Required: Sterility by Membrane Filtration Steritest
Date Received: 11/05/2016
Test Article: TEGO 2000sc
Sample Code: FMP16111 40355
Batch Ref: 530251
Qty Received: 20 x 50mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Growth of Sub-culture in TSB at 20-25°C after 4 - 7 days	MM107/07	No growth in two broths
Growth of Sub-culture in THY at 30-35°C after 4 - 7 days	MM107/08	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 8.0 2.6.1, USP 39 <71> & JP XVI 4.06
Product Standard Data Sheet	FG047/psd	FM08

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 09 Jun 2016 14:45:45

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology