



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 18.06.2016

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	FMP 16139 40411
Production Date	18.05.16
Expiration Date	EXP 18/03/2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	clear Slightly yellow liquid
pH 100% (20°C)	JDM001	8.0	9.0	8.9
Specific Gravity (20°C)	JDM004	0.980	1.020	1.003

On behalf of Diversey site Quality Manager	Name :	J. Staan, P. Tatyish
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513403	Version : 04	Date of issuing : April 19 th 2012
-------------	--------------	---



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 14-Jun-2016

UK32S11632211-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
 Unit 8, White Cross Ind Estate
 Hightown
 Lancaster
 Lancashire LA1 4XS
 UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002758
Customer Reference Number:	P026686
Product Description:	TEGO 2001 SS 25-45kGy
Validation Reference:	4.1903
Quantity Received:	136
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	131 X FMP16139 40411 + 5 SAMPLES THUR 26TH MAY 2016 3 PLTS LOT 2

Irradiation Data

Date and Time of Irradiation:	11-Jun-2016 09:17
Reference Dose Range kGy:	36.3 - 36.9
Calculated Minimum Dose kGy:	25.7
Calculated Maximum Dose kGy:	38.6

Irradiation Release Authorised By Synergy Health plc

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069