



Diversey Europe Operations BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 10.08.2016

Product Name	CLEARKLENS CLEANSINALD SS VH9S
Product Code	100848254
Batch Number	FMP16223 - 41029
Production Date	10/08/2016
Expiration Date	EXP 10/02/2018

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	Clear slightly yellow liquid
pH (100% 20°C)	JDM001	9.0	11.4	11.1
Specific Gravity (20°C)	JDM004	0.990	1.010	1.000

On behalf of Diversey site Quality Manager	Name :	H. H. H. J. Staan
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 100848254	Version : 02	Date of issuing : April 7 th 2016
---------------	--------------	--

Certificate of Irradiation

Date Issued: 23-Jul-2016

UK33S11664986-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1106548
Customer Reference Number:	P026899
Product Description:	C/SINALDSS BLK3 DV4673 25-40kGy
Validation Reference:	4673
Quantity Received:	5
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	40.0
Customer Unit Lot/Batch Number:	DEV001906, WO41030, 1 plt

Irradiation Data

Date and Time of Irradiation:	23-Jul-2016 08:35
Reference Dose Range kGy:	32.8 - 33.0
Calculated Minimum Dose kGy:	30.1
Calculated Maximum Dose kGy:	36.1

Item irradiated under WO41030 will be used in finished batch code FMP16223 41029

Irradiation Release Authorised By Synergy Health plc

Processing Site: Brunel Close, Drayton Fields Industrial Estate, Daventry, NN11 8RB Phone No: +44 (0) 1327 706111

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 12 Aug 2016
Date Tested: 23 Aug 2016
Date Test Completed: 06 Sep 2016
Purchase Order: 27146

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0030495/1
Test Required: Sterility by Membrane Filtration Steritest
Date Received: 12/08/2016
Test Article: ClearKlens Cleansinald SS
Sample Code: FMP16223 41029
Batch Ref: 7822
Qty Received: 20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	See external comments
Product Standard Data Sheet	FG047/psd	FM06

Comments

The testing procedure was performed in compliance with the current Pharmacopoeias however, the volume tested does not comply with the requirements of the current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 07 Sep 2016 15:46:07

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology