

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 27.10.16

| Product Name | CLEARKLENS TEGO 2001SS VH24S | | |
|------------------------|------------------------------|--|--|
| Product Code | 7513403 | | |
| Batch Number | FMP 16179 40913 | | |
| Production Date | 27.06,206 | | |
| Expiration Date | EXP 271041 2017 | | |

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

| Test | Test Method | Limits | | Results |
|-------------------------|-------------|------------|-------------|----------------|
| | | Lower - | Upper | |
| Appearance | Visual | Clear Slig | htly Yellow | Clear Slightly |
| | | Liquid | | Yellon Lique |
| pH 100% (20°C) | JDM001 | 8.0 | 9.0 | 8.5 |
| Specific Gravity (20°C) | JDM004 | 0.980 | 1.020 | 1.000 |
| | | | | |

| On behalf of Diversey site Quality Manager | Name : | Parlyish M. Staron |
|---|----------|---------------------------|
| | Position | Quality Control Inspector |

| This document being issued electronical | ly does not bear a signature | |
|---|------------------------------|----------------------------------|
| COA 7513403 | Version: 04 | Date of issuing: April 19th 2012 |



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 17-Oct-2016 UK33S11660538-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information 100432 **Account Number: Synergy Health Sales Part Reference:** 1106485 P026863 **Customer Reference Number:** TEGO 2001SS 25-45kGy TRIPLE VALIDATION **Product Description: Ouantity Received:** 117 25.0 **Customer Minimum Specification kGy:** 45.0 **Customer Maximum Specification kGy:** VALIDATION, DEVO01156, FMP16179 40913, 2 plts Customer Unit Lot/Batch Number: Actual min dose: 29.3kGy Actual max dose: 42.4kGy **Other Process Details:** Irradiation Data 14-Oct-2016 13:35 Date and Time of Irradiation: