

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date:

3.03.2017

CLEARKLENS TEGO 2001SC VH24	
7517177	
FMP16328 41663	
23 11 2016	
EXP 23/11/2018	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test Method	Limits		Results
	Lower -	Upper	
Visual	Clear Sligi	ntly Yellow	Clear Blightly
	Lia	uid	Yellow Liquia
JDM001	8.5	9.5	8.6
JDM004	1.040	1.080	1,059
	Visual JDM001	Visual Clear Sligl Liq JDM001 8.5	Visual Clear Slightly Yellow Liquid JDM001 8.5 9.5

On behalf of Diversey site	Name:	Parly isto 1 Stars
Quality Manager	Position	Quality Control Inspector

<u> Inis aocument being issued electronically</u>	y does not bear a signature	
COA 7517177	72 1 0.0	Date of issuing: November 28th 2011





http://www.synergyhealthpic.com

Certificate of Irradiation

Date Issued: 30-Jan-2017

UK33S11767960-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information **Account Number:** 100432 Synergy Health Sales Part Reference: 1109782 **Customer Reference Number:** 27896 **Product Description:** TEGO 2001 SC SOLUTION 100ML 25-45kGy VALIDATION **Ouantity Received:** 27 Customer Unit Lot/Batch Number: FMP16328 41663, INC 1 SAMPLE, VALIDATION, 2 Other Process Details: Actual min dose: 28.0kGy Actual max dose: 39.8kGy **Irradiation Data** Date and Time of Irradiation: 27-Jan-2017 12:53

Irradiation Release Authorised By Synergy Health plc

A2



http://www.synergynealthpic.com

Certificate of Irradiation

Date Issued: 09-Feb-2017

UK33S11806934-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information				
Account Number:	100432			
Synergy Health Sales Part Reference:	1109782			
Customer Reference Number:	27896			
Product Description:	TEGO 2001 SC SOLUTION 100ML DV4856			
	25-45kGy			
Validation Reference:	4856			
Quantity Received:	74			
Customer Minimum Specification kGy:	25.0			
Customer Maximum Specification kGy:	45.0			
Customer Unit Lot/Batch Number:	FMP16328 41663, INC 1 SAMPLE			
lı	rradiation Data			
Date and Time of Irradiation:	09-Feb-2017 20:14			
Reference Dose Range kGy:	33.8 - 35.7			
Calculated Minimum Dose kGy:	26.7			
Calculated Maximum Dose kGy:	39.2			

Irradiation Release Authorised By Synergy Health plc