

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 13,05=17

| Product Name | CLEARKLENS TEGO 2001SS VH24S | | | |
|-----------------|------------------------------|--|--|--|
| Product Code | 7513403 | | | |
| Batch Number | FMP17112 42557 | | | |
| Production Date | 22/04/2017 | | | |
| Expiration Date | EXP 22/02/2018 | | | |

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

| Test | Test Method | Limits | | Results |
|-------------------------|-------------|-----------------------|-------|----------------|
| <u></u> | | Lower - | Upper | |
| Appearance | Visual | Clear Slightly Yellow | | clear slightly |
| | | Liquid | | Yellow Liquel |
| pH 100% (20°C) | JDM001 | 8.0 | 9.0 | 8.8 |
| Specific Gravity (20°C) | JDM004 | 0.980 | 1.020 | 1.000 |
| | | | | |

| | 0 1 1 10 001 | Name : | Makine Partyin | Sho |
|---|--------------|---------------------------|----------------|-----|
| On behalf of Diversey site Quality Manager | Position | Quality Control Inspector | | |

This document being issued electronically does not bear a signature

COA 7513403 Version: 04 Date of issuing: April 19th 2012



http://www.synergynealthpic.com

Certificate of Irradiation

Date Issued: 05-May-2017

UK33S11867319-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterllisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information **Account Number:** 100432 Synergy Health Sales Part Reference: 1106485 **Customer Reference Number:** P028853 **Product Description:** TEGO 2001SS DV4767 25-45kGy Validation Reference: 4767 **Quantity Received:** 134 Customer Minimum Specification kGv: 25.0 **Customer Maximum Specification kGy:** 45.0 Customer Unit Lot/Batch Number: FMP17112, 42557, INC 1 SAMPLE BOX, 3 PLT **Irradiation Data** Date and Time of Irradiation: 05-May-2017 02:35 Reference Dose Range kGy: 34.7 - 36.8 27.6 Calculated Minimum Dose kGy: Calculated Maximum Dose kGy: 41.7