

Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England Telephone: +44(0)1329 226600

Fax:

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

12 Sep 2017

Date Tested:

19 Sep 2017

Date Test Completed:

03 Oct 2017

**Purchase Order:** 

26905

### **CERTIFICATE OF ANALYSIS**

Laboratory Reference Number:

0037891/2

Test Required:

Sterility by Membrane Filtration Steritest

Date Received: Test Article: Sample Code:

12/09/2017

20 x 100mL

570742

Bi-Spore (Base) FMP17248 43365

**Batch Ref: Qty Received:** 

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 9.0 2.6.1, USP 40 <71> & JP XVII 4.06
Product Standard Data Sheet	FG047/psd	FM05

Approval is provided by Electronic Signature. Their name and position is shown below.

CBio\* MMSB

Date: 04 Oct 2017 12:53:37

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Certificate of Analysis - OSMM Consignment: 0037891 Print Number: P0046083 Page 1 of 1





Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England Telephone: +44(0)1329 226600

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Hightown

White Cross Industrial Estate

Lancaster LAT 4XS

Date Received: 12 Sep 2017

Fax:

Date Tested: 19 Sep 2017

Date Test Completed: 03 Oct 2017

Purchase Order: 26905

## **CERTIFICATE OF ANALYSIS**

Laboratory Reference Number:

0037891/1

Test Required:

Sterility by Membrane Filtration Steritest

Date Received:

12/09/2017

**Test Article:** Sample Code: Bi-Spore (Activator) FMP17248 43365

Batch Ref: 583158 **Qty Received:** 20 x 100mL

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 9.0 2.6.1, USP 40 <71> & JP XVII 4.06
Product Standard Data Sheet	FG047/psd	FM04

Approval is provided by Electronic Signature. Their name and position is shown below.

CBiol, MIXSB

Date: 04 Oct 2017 12:53:37

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Certificate of Analysis - OSMM Consignment: 0037891 Print Number: P0046083

Page 1 of 1







http://www.synergyhealthplc.com

# **Certificate of Irradiation**

Date Issued: 23-Aug-2017

UK33S11942898-3-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices


Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

#### Order Information **Account Number:** 100432 Synergy Health Sales Part Reference: 1106825 **Customer Reference Number:** P029431 **Product Description:** BISP B3 COMPONENTS DV4725 25-40kGv Validation Reference: 4725 **Quantity Received:** 8 Customer Minimum Specification kGy: 25.0 **Customer Maximum Specification kGy:** 40.0 Customer Unit Lot/Batch Number: W043363, 1PLT irradiation Data Date and Time of Irradiation: 22-Aug-2017 22:31 Reference Dose Range kGy: 33.4 - 34.8 Calculated Minimum Dose kGy: 27.7 Calculated Maximum Dose kGy: 35.5

Items irradiated under WO43363, will be used in finished product batch FMP17248 43365

Irradiation Release Authorised By Synergy Health plc





**Diversey Europe Operations BV** 

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

## **CERTIFICATE OF ANALYSIS**

Date: 05.09.2017

<b>Product Name</b>	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26		
<b>Product Code</b>	7514852		
<b>Batch Number</b>	FMP 17248 43365		
<b>Production Date</b>	05/09/2017		
<b>Expiration Date</b>	EKP 0510912019		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lir	nits	Results
		Lower -	Upper	
Appearance	Visual	Clear C	olourless	Clear Colour-
		Lic	quid	less Liquid
pH (100% 20°C)	JDM001	9.0	12.5	11.9
Specific Gravity (20°C)	JDM004	1.004	1.020	1.008
Potential Chlorine		100	180	110 =
Dioxide				148.5

0 1 1 10 00	Name:	Partyist Ogadinense
On behalf of Diversey site  Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7514852 Version: 04 Date of issuing: July 6th 2012



**Diversey Europe Operations BV** 

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

# **CERTIFICATE OF ANALYSIS**

Date: 05.09. 2017

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Product Code	7514852
Batch Number	FMP 17248 43365
Production Date	05/09/2017
	EXP 05/09/2019
Expiration Date	ENP OS OF IZO

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin Lower -	nits Upper	Results
Appearance	Visual		olourless Juid	clear Colour
pH (100% 20°C)	JDM001	1.5	2.5	1.6
Specific Gravity (20°C)	JDM004	1.010	1.030	1.020

The state of the s	Name:	Party isto apaderonno.
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector

This document being issued	electronically does not bear a signature	C: 11 4h 2012
COA 7514852	Version: 04	Date of issuing : July 6th 2012