



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 09/04/2018

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	FMP 18073 44720
Production Date	14/03/2018
Expiration Date	EXP 14/01/2019

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	clear slightly yellow liquid
pH (neat solution)	DM001	8.0	9.0	8.6
Specific Gravity (20°C)	DM004	0.980	1.020	1.000

On behalf of Diversey site Quality Manager	Name :	Pantjeiska Karabrode
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24 th 2017
--------------	--------------	--



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 28-Mar-2018

UK33S12080357-2-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1106485
Customer Reference Number:	P030700
Product Description:	TEGO 2001SS DV4767 25-45kGy
Validation Reference:	4767
Quantity Received:	134
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	FMP18073 44720, INC 1 SAMPLE, 3 plts

Irradiation Data

Date and Time of Irradiation:	28-Mar-2018 12:26
Reference Dose Range kGy:	35.6 - 36.2
Calculated Minimum Dose kGy:	28.3
Calculated Maximum Dose kGy:	41.0

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company

Processing Site: Brunel Close, Drayton Fields Industrial Estate, Daventry, NN11 8RB Phone No: +44 (0) 1327 706111

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069