



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 10.06.2019

Product Name	CLEARKLENS DE VH29
Product Code	100882174
Batch Number	7MP19115 47174
Production Date	25.04.2019
Expiration Date	EXP 26/04/2021

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear, colourless liquid
Specific Gravity (20°C)	DM004	0.880	0.888	0.885

On behalf of Diversey site Quality Manager	Name :	J. Stanon, Parbyes
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24 th 2017
--------------	--------------	--



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hosford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 06 Jun 2019
Date Tested: 09 Jul 2019
Date Test Completed: 23 Jul 2019
Purchase Order: 33232

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0049051/5
Test Required: Sterility by Membrane Filtration Steritest
Date Received: 06/06/2019
Test Article: Clearklens DE
Sample Code: FMP19115 47174
Batch Ref: 10227
Qty Received: 20 x 900mL Bottles + 2 Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	See external comments
Product Standard Data Sheet	FG047/psd	FM03-02

Comments

The testing procedure was performed in compliance with the current Pharmacopoeias however, the volume tested does not comply with the requirements of the current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

 BSc. CBiol. MRSA

Date: 24 Jul 2019 13:01:50

Mrs P. Pham-Lengoc
Documentation Manager – Microbiology



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 06 Jun 2019
Date Tested: 07 Jun 2019
Date Test Completed: 07 Jun 2019
Purchase Order: 33232

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0049051/6
Test Required: Bacterial Endotoxin Testing by Kinetic QCL
Date Received: 06/06/2019
Test Article: Clearklens DE
Sample Code: FMP19115 47174
Batch Ref: 10227
Qty Received: 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL
Spiked Recovery	MM110/01	128 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 9.3 2.6.14, USP 42 <85> & JPXVII 4.01
Product Standard Data Sheet	FG047/psd	FM01-02

Approval is provided by Electronic Signature. Their name and position is shown below.



Date: 12 Jun 2019 14:13:30

Victoria Watson
Laboratory Manager – Microbiology

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID: 2173-3389A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
DIVER DE BLK2 BOTTLE DV4648	FMP19115 47174	109	Case
Validation Reference Number: 4648			

Processing Run Start Date: 28-May-19 03:12 AM

Processing Run End Date: 28-May-19 11:35 AM

Specified Dose Range (kGy):	25.0 - 45.0	Calculated Min Dose (kGy):	27.5
Reference Dose Range (kGy):	31.6- 39.8	Calculated Max Dose (kGy):	40.9

Gamma Process Run Approval authorized by STERIS

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK
Limited, a STERIS Company
Brunel Close
Drayton Fields Industrial Estate
Daventry
Northants
NN11 8RB
Phone: + 44(0) 1327 706 111

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID: 2173-3288A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
DIVER DE BLK2 BOTTLE DV4648	FMP19115 47174	486	Case
Validation Reference Number: 4648			
DIVER DE BLK2 BOTTLE DV4648	FMP19115 47174 Inc 2 samples	56	Case
Validation Reference Number: 4648			
DIVER DE BLK2 BOTTLE DV4648	FMP19115 47174 Inc 5 samples	59	Case
Validation Reference Number: 4648			

Processing Run Start Date: 28-May-19 12:09 AM

Processing Run End Date: 28-May-19 10:30 AM

Specified Dose Range (kGy):	25.0 - 45.0	Calculated Min Dose (kGy):	27.5
Reference Dose Range (kGy):	31.6- 39.8	Calculated Max Dose (kGy):	41.0

Gamma Process Run Approval authorized by STERIS

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK
Limited, a STERIS Company
Brunel Close
Drayton Fields Industrial Estate
Daventry
Northants
NN11 8RB
Phone: + 44(0) 1327 706 111