

versey Europe Operations BV aarssenbroeksedijk 2 42 DN Utrecht ie Netherlands

I. +31 (0)30 247 6427 ww.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 14 November 2019

Product Name	CLEA	CLEARKLENS DE VH29		
Product Code	101101081			
Batch Number	FMP 19301	48637		
Production Date	28/10/2019			
Expiration Date	EXP 28/10/2021			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits Lower - Upper		Results
Appearance	Visual	Clear Color	urless Liquid	lear colourless Liqui
Specific Gravity (20°C)	DM004	0.880	0.888	0.885

On behalf of Diversey site	Name:	urszula Haraburda Angelíka Partynska
Quality Manger	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

1		_ , ., ., ., ., ., ., ., ., ., ., ., ., .
COA Template	Version: 02	Date of issuina : November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 11/11/2019 9:34:46 GMT

Page 1 of 1

Cianad Die

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2551722 P.O.# PO34216 Sales Order # 2364985

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 11/07/2019 03:37:30 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	483 BX	100862174	Diversey Clearklens DE 900 mL	FMP19296	WO48482/DEV001910
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	374 BX	101101081	Diversey Clearklens DE 900ml	FMP19301	WO48637/DEV002096
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	3 BX	101101081_SAMPLE	Diversey Clearklens DE 900ml	FMP19301	WO48637/DEV002096
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	860 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.3 KGY	Pass	MFORD MARTYN FORD	11/07/2019 17:07:42 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy t	39.1 KGY	Pass	MFORD MARTYN FORD	11/07/2019 17:07:57 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 11/11/2019 09:33:51 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

11/15/2019 12:38:00 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2554435 P.O.# PO34262 Sales Order # 2367944

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 11/14/2019 09:13:54 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	62 BX	100862174	Diversey Clearklens DE 900 mL	FMP19296	WO48482/DEV001910
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	335 BX	101101081	Diversey Clearklens DE 900ml	FMP19301	WO48637/DEV002096
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	380 BX	7513400	Diversey Clearklens IPA 900 mL	FMP19311	WO48699/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
104.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP19311	WO48699/DEV001152
	Dose Map 780 BX	302_0030 Total	Ref Pos: 10E1, AF Min = 0.71		

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	26.4 KGY	Pass	IWEATHERAL	11/15/2019 03:57:43 GMT
		Reason Code Test				IAN WEATHERALL	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	38.9 KGY	Pass	IWEATHERAL	11/15/2019 03:58:07 GMT
		Reason Code Test				IAN WEATHERALL	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON Date: 11/15/2019 12:36:46 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens DE

Sample Description:

Sample Code: Fmp19301 48637

Lucideon Sample Number: (196270)-47834

Lucideon Report Number: (196270)-47834/ETEP

Issue Number:

Date Logged:

16-Nov-2019

Order Number:

34301

Date Reported:

26-Nov-2019

Date(s) of Test(s):

18-Nov-2019 to 18-Nov-2019

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens DE

Sample Description:

Sample Code: Fmp19301 48637

Lucideon Sample Number: (196270)-47833

Lucideon Report Number: (196270)-47833/MFEP

Issue Number:

1

Date Logged:

16-Nov-2019

Order Number:

34301

Date Reported:

05-Dec-2019

Date(s) of Test(s):

17-Nov-2019 to 01-Dec-2019

Sterility Testing Membrane Filtration EP

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass

Tests carried out in accordance with cGMP

End of Test Report

N bBoot as -Dec-19

Mrs Natalie Boot

Senior Business Support Assistant

Staffordshire ST4 7LQ