

Diversey Europe Opertaions BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

10.09.2019

Product Name	CLEARKLENS CLEANSINALD SC VH9		
Product Code	7516429		
Batch Number	TMP 19253 47642		
Production Date	10.03.2019		
Expiration Date	EXP 10/09/2021		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results	
		Lower -	Upper		
Appearance	Visual	Clear Slightly Yellow Liquid		clear slightly	
pH (neat solution)	DM001	12.0	13.0	U	3
Specific Gravity (20°C)	DM004	1.040	1.060	1.	055

On behalf of Diversey site Quality Manager	Name:	P. Stavan A Bonys
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing: November 24 2017
--------------	--------------	-----------------------------------

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID: 2173-4746A

Product Code Lot Number Quantity UOM

C/SINALD BOTTLES/CAPS/BAGS W047643

DV4894

Validation Reference Number: 4894

Processing Run Start Date: 25-Aug-19 06:46 PM

Processing Run End Date: 26-Aug-19 01:08 AM

Specified Dose Range (kGy): 25.0 - 40.0 Calculated Min Dose (kGy): 29.9

Reference Dose Range (kGy): 27.4- 36.5 Calculated Max Dose (kGy): 36.4

Gamma Process Run Approval authorized by STERIS

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK Limited, a STERIS Company Brunel Close Drayton Fields Industrial Estate Daventry Northants NN11 8RB

Phone: +44(0) 1327 706 111

Items irradiated under WO47643 will be used in finished batch FMP19253 47642

Olga Kirchner

Rel Date: 08/13/2018

11

Case

Document ID: 27387

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

Mrs.Olga Kirchner FAO:

Report of Tests on:

Cleansinald SC

Sample Description:

Sample Code: FMP19253 47642

Lucideon Sample Number: (195132)-39887

Lucideon Report Number: (195132)-39887/MFEP

Issue Number:

Date Logged:

16-Sep-2019

Order Number:

PO 33882

Date Reported:

03-Jan-2020

Date(s) of Test(s):

21-Nov-2019 to 03-Jan-2020

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

