

Diversey Europe Operations BV Magrasenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 03.02 2020

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7515828
Batch Number	FNP20034, 49138
Production Date	05.02.2020
Expiration Date	ExP03/02/2012

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	_	nits - Upper	Results	
Appearance	Visual			cleary colourless !	DUC
pH (neat solution)	DM001	9.0	12.5	11.4	1
Specific Gravity (20°C)	DM004	1.004	1.020	1.012	
Potential Chlorine Dioxide (ppm)		100	180	121.5 ppm	

On behalf of Diversey site	Name:	P.Stavar A. Borgs
Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version: 02	Date of issuing: November 24 2017
--------------	-------------	-----------------------------------



Diversey Europe BV Maarasenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 03. 02 2020

CLEARKLENS BI-SPORE BASE SOLUTION VI		
7515828		
FHP 20034, 49138		
03.02.2020		
EXP2/2012		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
		Lower - Upper	
Appearance	Visual	Clear Colourless Liquid	clear adamless qu
pH (neat solution)	DM001	1.5 2.5	2.
Specific Gravity (20°C)	DM004	1.010 1.030	1.024

On behalf of Diversey site	Name:	Y. Stavan A. Bonys
Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing: November 24th 2017
--------------	--------------	-------------------------------------

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID: 2173-7354A

Product CodeLot NumberQuantityUOMBI-SP B3 COMPONENTS DV4725WO491378Case

Validation Reference Number: 4725

Processing Run Start Date: 10-Jan-20 04:40 PM
Processing Run End Date: 11-Jan-20 01:18 AM

Specified Dose Range (kGy): 25.0 - 40.0 Calculated Min Dose (kGy): 27.3 Reference Dose Range (kGy): 30.2- 39.2 Calculated Max Dose (kGy): 34.1

PO Number: PO34552 MOC APPLIES TO THIS ORDER

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest

Reviewed and E-Signed By: Bhavya Ramisetty (Quality Engineer)

Date/Time E-Signed: 2020-01-23 08:28 AM

Document Content Revision: 1

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK Limited, a STERIS Company Brunel Close Drayton Fields Industrial Estate Daventry Northants NN11 8RB

Phone: + 44(0) 1327 706 111

Items irradiated under WO49137 will be used in finished batch Fmp20034 49138.

Olga Kirchner

Document ID: 40873

Rel Date: 08/13/2018

Last Revised in Rel 2.0.0.0

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

Mrs.Olga Kirchner FAO:

Report of Tests on:

Bi-Spore Activator

Sample Description:

Sample Code: Fmp20034 49138

Lucideon Sample Number: UK20745-5570

Lucideon Report Number: UK20745-5570/MFEP

Issue Number:

1

Date Logged:

13-Feb-2020

Order Number:

PO 34767

Date Reported:

05-Mar-2020

Date(s) of Test(s):

18-Feb-2020 to 03-Mar-2020

Sterility Testing Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report



Mrs Natalie Boot

Senior Business Support Assistant

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Base

Sample Description:

Sample Code: Fmp20034 49138

Lucideon Sample Number: UK20745-5571

Lucideon Report Number: UK20745-5571/MFEP

Issue Number:

Date Logged:

13-Feb-2020

Order Number:

PO 34767

Date Reported:

05-Mar-2020

Date(s) of Test(s):

18-Feb-2020 to 03-Mar-2020

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

