

Certificate of Analysis

Date 28/05/2020

Article
SKU 100934589 ClearKlens oxifast 1L sterile

page 1/1

batch: SEP/1547804/20/091
expiry date: 2022-03
date of manufacture: 30/03/2020

Characteristic Method	Unit	Value	lower limit	upper limit	conformity
peracetic acid MAN LCQ 011	%	0,08	0,07	0,10	Yes
hydrogen peroxide MAN LCQ 011	%	3,00	2,90	3,40	Yes
loss TAO 16h at 96°C MAN LCQ 009	%	0,00	0,00	2,50	Yes
Sterility certificate European Pharmacopeia 2.6.1					Yes

released on : 28/05/2020

by : LALEVEE

Comments:

Sterility certificate n° 200508-0076-001

The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the MSDS.

End of certificate of analysis