

versey Europe Operations BV aarasenbroeksedijk 2 42 DN Utrecht a Netherlands

il. +31 (0)30 247 6427 ww.Diversey.com

### **CERTIFICATE OF ANALYSIS**

Date:

03 July 2020

Product Name	CLEARKLENS IPA VH1				
Product Code	7513400				
Batch Number	FMP20154	50228			
Production Date	02/06/2020				
Expiration Date	EXP 02/06/2022				

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	its - Upper	Results  Clear Colourless Liquid  0.873	
Appearance	Visual	Clear Colou	ırless Liquid		
Specific Gravity (20°C)	DM004	0.872	0.883		

On behalf of Diversey site	Name:	Justyna Staron	Urszula Haraburda
Quality Manger	Position	Quality	Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



## **Certificate of Processing**

# STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 06/2

06/26/2020 10:41:50 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2685331 P.O.# PO35683 Sales Order # 2524361

**25 - 45 kGy** Gamma Irradiation Irradiation Date/Time: 06/25/2020 12:31:18 GMT

SO Line#	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	783 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20154	WO50228/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20154	WO50228/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	54 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20161	WO49631/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	840 BX	Total			

#### **Quality Test Summary**

		Signed By					
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.4 KGY	Pass	AWAGSTAFF ADDISON WAGSTAFF	06/26/2020 02:51:46 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.8 KGY	Pass	ESILLS EMMA SILLS	06/26/2020 10:38:59 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 06/26/2020 10:39:51 GMT

Reason: Work Order Completions



## **Certificate of Processing**

# STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 06

06/26/2020 13:41:34 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2681973 P.O.# PO35634 Sales Order # 2520137

**25 - 45 kGy** Gamma Irradiation Irradiation Date/Time: 06/25/2020 21:40:42 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	204 BX	101101081	Diversey Clearklens DE 900ml	FMP20149	WO49601/DEV002096
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	10 BX	101101081_SAMPLE	Diversey Clearklens DE 900ml	FMP20149	WO49601/DEV002096
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	646 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20154	WO50228/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	860 BX	Total			

#### **Quality Test Summary**

		Signed By					
Ор#	<b>Quality Test Description</b>	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.5 KGY	Pass	SWHELAN STEVEN WHELAN	06/26/2020 11:00:06 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.8 KGY	Pass	SWHELAN STEVEN WHELAN	06/26/2020 11:00:30 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 06/26/2020 13:39:36 GMT

Reason: Work Order Completions

### PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO: Mrs.Olga Kirchner

**Sample Description:** 

Clear Klens IPA Report of Tests on:

Lucideon Sample Number: UK202739-19926

Lucideon Report Number: UK202739-19926/MFEP Issue Number:

Date Logged: Order Number: 06-Jul-2020 PO 35858

Sample Code: Fmp20154 50228

Date(s) of Test(s): **Date Reported:** 27-Jul-2020 09-Jul-2020 to 23-Jul-2020

**Sterility Testing** 

Membrane Filtration EP


**Test Results:** 

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report** 



Mrs Andrea Saunders

Lucideon Limited

Stoke-on-Trent

## PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20154 50228

Lucideon Sample Number: UK202739-19927

Lucideon Report Number: UK202739-19927/ETEP

Issue Number:

Date Logged:

06-Jul-2020

Order Number:

PO 35858

**Date Reported:** 

Date(s) of Test(s):

30-Jul-2020

to 30-Jul-2020

30-Jul-2020

**Endotoxin Test (EP)** 

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

**End of Test Report** 

Mrs Andrea Saunders

**Business Support Assistant** 

Page 1 of 1