



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 08 July 2020

Product Name	CLEARKLENS IPA VH1		
Product Code	7513400		
Batch Number	FMP20161	49631	
Production Date	09/06/2020		
Expiration Date	EXP 09/06/2022		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	- Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manger	Name:	Angelika Partynska Urszula Haraburda
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24th
--------------	--------------	---------------------------------

R55480102

 06/26/2020 10:41:50 GMT
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2685331
P.O.#	PO35683			Sales Order #	2524361
	25 - 45 kGy	Gamma Irradiation		Irradiation Date/Time:	06/25/2020 12:31:18 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	783	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20154	WO50228/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20154	WO50228/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	54	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20161	WO49631/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	840	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.4 KGY	Pass	AWAGSTAFF	06/26/2020 02:51:46 GMT
		Reason Code Test				ADDISON WAGSTAFF	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.8 KGY	Pass	ESILLS	06/26/2020 10:38:59 GMT
		Reason Code Test				EMMA SILLS	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
 Reason: Work Order Completions

Date: 06/26/2020 10:39:51 GMT

R55480102

07/06/2020 9:13:53 GMT
Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2686119
P.O.#	PO35696			Sales Order #	2525455
25 - 45 kGy		Gamma Irradiation		Irradiation Date/Time:	07/03/2020 16:35:02 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	860	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20161	WO49631/DEV001152
			Dose Map	302_0030		
	860	BX	Total	Ref Pos: 10E1, AF Min = 0.71		

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By-----	
						User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.0 KGY	Pass	DMARRIOTT	07/04/2020 04:45:40 GMT
		Reason Code Test				DAVID MARRIOTT	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	41.0 KGY	Pass	DMARRIOTT	07/04/2020 04:46:03 GMT
		Reason Code Test				DAVID MARRIOTT	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON
Reason: Work Order Completions

Date: 07/06/2020 09:11:41 GMT

R55480102

 07/07/2020 9:13:45 GMT
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2687940
P.O.#	PO35743			Sales Order #	2527566
	25 - 45 kGy	Gamma Irradiation	Irradiation Date/Time:	07/06/2020 15:36:10 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	515	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20161	WO49631/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20161	WO49631/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	322	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20164	WO49640/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	840	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Maximum Dose	25.0 kGy	45.0 kGy	41.6 KGY	Pass	AHOBSON	07/07/2020 09:03:04 GMT
		Reason Code Test				ASHLEY HOBSON	
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.1 KGY	Pass	BCZAPLA	07/07/2020 03:51:13 GMT
		Reason Code Test				BARTOSZ CZAPLA	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON
 Reason: Work Order Completions

Date: 07/07/2020 09:11:26 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: Fmp20161 49631

Lucideon Sample Number: UK202816-20268

Lucideon Report Number: UK202816-20268/ETEP **Issue Number:** 1

Date Logged: 10-Jul-2020 **Order Number:** PO35888

Date Reported: 30-Jul-2020 **Date(s) of Test(s):** 30-Jul-2020 to 30-Jul-2020

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report



30 JUL 20

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: Fmp20161 49631

Lucideon Sample Number: UK202816-20267

Lucideon Report Number: UK202816-20267/MFEP **Issue Number:** 1

Date Logged: 10-Jul-2020 **Order Number:** PO35888

Date Reported: 31-Jul-2020 **Date(s) of Test(s):** 17-Jul-2020 to 31-Jul-2020

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report


31 Jul 20

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com