



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 04 August 2020

Product Name	CLEARKLENS TEGO 2000SC VH25		
Product Code	100868202		
Batch Number	FMP20202	50113	
Production Date	20/07/2020		
Expiration Date	EXP 20/07/2023		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid		Clear Slightly Yellow Liquid
Specific Gravity (20°C)	DM004	0.993	1.003	1.002
pH (neat solution)	DM001	7.5	8.5	8.1

On behalf of Diversey site Quality Manger	Name:	Angelika Partynska Urszula Haraburda
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24th
--------------	--------------	---------------------------------

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID: 2173-12305A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
TEGO 2000SC DV4724	FMP20202 WO50113	50	Case
Validation Reference Number: 4724			
TEGO 2000SC DV4724	FMP20202 WO50113, INC 1 SMPL	40	Case
Validation Reference Number: 4724			

Processing Run Start Date: 28-Jul-20 08:54 AM

Processing Run End Date: 28-Jul-20 05:55 PM

Specified Dose Range (kGy):	25.0 - 45.0	Calculated Min Dose (kGy):	29.8
Reference Dose Range (kGy):	29.4- 40.5	Calculated Max Dose (kGy):	40.3

PO Number: 35983

Gamma Process Run Approval authorized by STERIS

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK
Limited, a STERIS Company
Brunel Close
Drayton Fields Industrial Estate
Daventry
Northants
NN11 8RB
Phone: + 44(0) 1327 706 111

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Tego 2000SC

Sample Description: Sample Code: Fmp20202 50113

Lucideon Sample Number: UK203165-22406

Lucideon Report Number: UK203165-22406/MFEP **Issue Number:** 1

Date Logged: 05-Aug-2020 **Order Number:** PO36053

Date Reported: 29-Sep-2020 **Date(s) of Test(s):** 20-Aug-2020 to 25-Sep-2020

Sterility Testing

Membrane Filtration EP

Test Results:


Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

Refer to deviation report DEV398.

End of Test Report


29 SEP 20

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com