



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 16 September 2020

Product Name	CLEARKLENS IPA VH1		
Product Code	7513400		
Batch Number	FMP20233	50030	
Production Date	20/08/2020		
Expiration Date	EXP 20/08/2022		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manger	Name:	Angelika Partynska Urszula Haraburda
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24th
--------------	--------------	---------------------------------

R55480102

 09/01/2020 9:44:05 GMT
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2726876
P.O.#	PO36258			Sales Order #	2575054
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	08/30/2020 07:28:35 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	840	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20233	WO50030/DEV001152
			Dose Map	302_0030		
	840	BX	Total	Ref Pos: 10E1, AF Min = 0.71		

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By-----	
						User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	28.3 KGY	Pass	JYATES	08/31/2020 05:06:30 GMT
		Reason Code Test				JOHN YATES	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	42.7 KGY	Pass	AHOBSON	09/01/2020 09:31:58 GMT
		Reason Code Test				ASHLEY HOBSON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON
 Reason: Work Order Completions

Date: 09/01/2020 09:41:37 GMT

R55480102

09/07/2020 8:56:55 GMT
Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	2729214
P.O.#	PO36291			Sales Order #	2577969
25 - 45 kGy		Gamma Irradiation		Irradiation Date/Time:	09/04/2020 18:45:09 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	577	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20233	WO50030/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20233	WO50030/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	580	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	28.8 KGY	Pass	SWHELAN	09/07/2020 08:12:41 GMT
		Reason Code Test				STEVEN WHELAN	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	41.7 KGY	Pass	AWAGSTAFF	09/05/2020 14:26:46 GMT
		Reason Code Test				ADDISON WAGSTAFF	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
Reason: Work Order Completions

Date: 09/07/2020 08:54:35 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: ClearKlens IPA

Sample Description: Sample Code: Fmp20233 50030

Lucideon Sample Number: UK203751-26384

Lucideon Report Number: UK203751-26384/MFEP **Issue Number:** 1

Date Logged: 15-Sep-2020 **Order Number:** PO36367

Date Reported: 09-Oct-2020 **Date(s) of Test(s):** 24-Sep-2020 to 08-Oct-2020

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report


09 OCT 20

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: ClearKlens IPA

Sample Description: Sample Code: Fmp20233 50030

Lucideon Sample Number: UK203751-26385

Lucideon Report Number: UK203751-26385/ETEP **Issue Number:** 1

Date Logged: 15-Sep-2020 **Order Number:** PO36367

Date Reported: 16-Oct-2020 **Date(s) of Test(s):** 01-Oct-2020 to 01-Oct-2020

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

Nc Boot 16-Oct-20

Mrs Natalie Boot

Senior Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com