



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 08 October 2020

Product Name	CLEARKLENS IPA VH1		
Product Code	7513400		
Batch Number	FMP20268	50665	
Production Date	24/09/2020		
Expiration Date	EXP 24/09/2022		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.872	0.883	0.873

On behalf of Diversey site Quality Manger	Name:	Angelika Partynska Urszula Haraburda
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24th
--------------	--------------	---------------------------------

R55480102

 10/05/2020 10:22:32 GMT
 Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER	Processing Facility: Markham Vale	Work Order # 2748090
P.O.# PO36484		Sales Order # 2600703
25 - 45 kGy	Gamma Irradiation	Irradiation Date/Time: 10/02/2020 15:31:45 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	118	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20266	WO50583/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20266	WO50583/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	352	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20267	WO50202/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
104.000	3	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20267	WO50202/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
105.000	351	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20268	WO50665/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
106.000	3	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20268	WO50665/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
107.000	130	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20269	WO50736/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	960	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.3 KGY	Pass	DMARRIOTT	10/03/2020 04:33:49 GMT
		Reason Code Test				DAVID MARRIOTT	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	40.7 KGY	Pass	DMARRIOTT	10/03/2020 04:34:13 GMT
		Reason Code Test				DAVID MARRIOTT	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

 Electronically Signed By: **EMMA SILLS**
 Reason: **Work Order Completions**

Date: 10/05/2020 10:20:01 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: Fmp20268 50665

Lucideon Sample Number: UK204118-28698

Lucideon Report Number: UK204118-28698/ETEP **Issue Number:** 1

Date Logged: 09-Oct-2020 **Order Number:** PO36590

Date Reported: 16-Oct-2020 **Date(s) of Test(s):** 15-Oct-2020 to 15-Oct-2020

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report


16 OCT 20

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Clear Klens IPA

Sample Description: Sample Code: Fmp20268 50665

Lucideon Sample Number: UK204118-28697

Lucideon Report Number: UK204118-28697/MFEP **Issue Number:** 1

Date Logged: 09-Oct-2020 **Order Number:** PO36590

Date Reported: 03-Nov-2020 **Date(s) of Test(s):** 16-Oct-2020 to 30-Oct-2020

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report


03NOV20

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com