

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

08 October 2020

Product Name	CLEARKLENS IPA VH1					
Product Code	7513400					
Batch Number	FMP20262	50033				
Production Date	18/09/2020					
Expiration Date	EXP 18/09/2022					

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results	
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid	
Specific Gravity (20°C)	DM004	0.872	0.883	0.873	

On behalf of Diversey site	Name:	Angelika Partynska Urszula Haraburda		
Quality Manger	Position	Quality Control Inspector		

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 09/2

09/28/2020 9:52:25 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2743616 P.O.# P036424 Sales Order # 2595214

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 09/26/2020 22:27:49 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	535 BX	101100418	Diversey Clearklens IPA/DE 900	FMP20260	WO50032/DEV002080
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	101100418_SAMPLE	Diversey Clearklens IPA/DE 900	FMP20260	WO50032/DEV002080
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	362 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20262	WO50033/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	900 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.7 KGY	Pass	DMARRIOTT DAVID MARRIOTT	09/27/2020 15:18:54 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	41.2 KGY	Pass	DMARRIOTT DAVID MARRIOTT	09/27/2020 15:19:13 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 09/28/2020 09:49:59 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

10/05/2020 9:38:59 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2745857 P.O.# PO36457 Sales Order # 2598106

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/02/2020 08:13:06 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	518	вх	7513400	Diversey Clearklens IPA 900 mL	FMP20262	WO50033/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20262	WO50033/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	6	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	NA	T020/120
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
104.000	413	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20266	WO50583/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	940	вх	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.9 KGY	Pass	DMARRIOTT	10/03/2020 04:37:07 GMT
		Reason Code Test				DAVID MARRIOTT	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	40.7 KGY	Pass	DMARRIOTT	10/03/2020 04:37:27 GMT
		Reason Code Test				DAVID MARRIOTT	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 10/05/2020 09:36:28 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20262 50033

Lucideon Sample Number: UK204118-28700

Lucideon Report Number: UK204118-28700/ETEP

Issue Number:

Date Logged:

09-Oct-2020

Order Number:

PO36590

Date Reported:

16-Oct-2020

Date(s) of Test(s):

15-Oct-2020 to 15-Oct-2020

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method


Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report



Mrs Andrea Saunders

Business Support Assistant

Lucideon Limited

Stoke-on-Trent

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20262 50033

Lucideon Sample Number: UK204118-28699

Lucideon Report Number: UK204118-28699/MFEP

Issue Number:

Date Logged:

09-Oct-2020

Order Number:

PO36590

Date Reported:

03-Nov-2020

Date(s) of Test(s):

16-Oct-2020 to 30-Oct-2020

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

Mrs Andrea Saunders

Business Support Assistant