

Diversey Europe Operations BV Maarasenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

23 October 2020

Product Name	CLEARKLENS IPA VH1			
Product Code	7513400			
Batch Number	FMP20279	50203		
Production Date	05/10/2020			
Expiration Date	EXP 05/10/2022			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results Clear Colourless Liquid	
Appearance	Visual	Clear Colou	rless Liquid		
Specific Gravity (20°C)	DM004	0.872 0.883		0.876	

On behalf of Diversey site Quality Manger	Name:	Justyna Staron Angelika Partynsk	
	Position	Quality	Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

10/19/2020 8:04:46 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2754150 P.O.# P036578 Sales Order # 2608211

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/18/2020 06:25:22 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	97 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20274	WO50737/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20274	WO50737/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	298 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20279	WO50203/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
104.000	2 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20279	WO50203/DEV001152
	Dose Map 400 BX	302_0030 Total	Ref Pos: 10E1, AF Min = 0.71		

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.8 KGY	Pass	BCZAPLA	10/18/2020 16:48:28 GMT
		Reason Code Test				BARTOSZ CZAPLA	
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.1 KGY	Pass	BCZAPLA	10/18/2020 16:48:42 GMT
		Reason Code Test				BARTOSZ CZAPLA	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 10/19/2020 08:02:11 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

10/19/2020 13:19:28 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2756852 P.O.# PO36602 Sales Order # 2611420

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/18/2020 01:11:11 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	54 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20279	WO50203/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20279	WO50203/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	903 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20281	WO50908/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	960 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	38.9 KGY	Pass	BCZAPLA BARTOSZ CZAPLA	10/18/2020 16:46:37 GMT
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	26.9 KGY	Pass	BCZAPLA BARTOSZ CZAPLA	10/18/2020 16:47:00 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 10/19/2020 13:16:40 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20279 50203

Lucideon Sample Number: UK204345-30585

Lucideon Report Number: UK204345-30585/ETEP

Issue Number:

Date Logged:

23-Oct-2020

Order Number:

PO36667

Date Reported:

16-Nov-2020

Date(s) of Test(s):

12-Nov-2020 to 12-Nov-2020

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method


Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report



Mrs Andrea Saunders

Business Support Assistant

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20279 50203

Lucideon Sample Number: UK204345-30584

Lucideon Report Number: UK204345-30584/MFEP

Issue Number:

Date Logged:

23-Oct-2020

Order Number:

PO36667

Date Reported:

27-Nov-2020

Date(s) of Test(s):

12-Nov-2020 to 26-Nov-2020

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report



Mrs Andrea Saunders

Business Support Assistant