

Diversey Europe Operations BV Hasrasenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

16 December 2020

Product Name	CLEARKLENS IPA VH1					
Product Code	7514372					
Batch Number	FMP20336	50309				
Production Date	01/12/2020					
Expiration Date	EXP 01/12/2022					

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method Limits Lower - Upper				
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid	
Specific Gravity (20°C)	DM004	0.872	0.883	0.875	

On behalf of Diversey site	Name:	urszula Haraburda Angelika Partynska
Quality Manger	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 12/15

12/15/2020 8:53:57 GMT Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2789176
P.O.# PO36959 Sales Order # 2650352

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 12/11/2020 18:29:14 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	872 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20331	WO50916/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20331	WO50916/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	65 BX	7514372	Diversey Clearklens IPA 900 mL	FMP20336	WO50309/DEV001250
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	940 BX	Total			

Quality Test Summary

			quality 10	or ourminar,	,	Signed By	-
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.3 KGY	Pass	SWARD SIMON WARD	12/12/2020 12:30:14 GMT
450.00	Maximum Dose	25.0 kGy	45.0 kGy	38.7 KGY	Pass	SWARD	12/12/2020 12:31:22 GMT
		Reason Code Test				SIMON WARD	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 12/15/2020 08:50:57 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

12/15/2020 8:50:00 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2790160 P.O.# PO36967 Sales Order # 2651482

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 12/11/2020 23:50:09 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	646	вх	7514372	Diversey Clearklens IPA 900 mL	FMP20336	WO50309/DEV001250
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	14	BX	7514372 Sample	Diversey Clearklens IPA 900 mL	FMP20336	WO50309/DEV001250
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	660	вх	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	26.8 KGY	Pass	SWARD SIMON WARD	12/12/2020 12:25:11 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.9 KGY	Pass	SWARD SIMON WARD	12/12/2020 12:25:38 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 12/15/2020 08:46:55 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20336 50309

Lucideon Sample Number: UK205168-36179

Lucideon Report Number: UK205168-36179/MFEP

Issue Number:

Date Logged:

17-Dec-2020

Order Number:

PO 37052

Date Reported:

22-Jan-2021

Date(s) of Test(s):

08-Jan-2021 to 22-Jan-2021

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

10ROOK 22-JON-21

Mrs Natalie Boot

Business Support Administrator

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20336 50309

Lucideon Sample Number: UK205168-36180

Lucideon Report Number: UK205168-36180/ETEP

Issue Number:

Date Logged:

17-Dec-2020

Order Number:

PO 37052

Date Reported:

22-Jan-2021

Date(s) of Test(s):

21-Jan-2021 to 21-Jan-2021

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

nRont 22-Jon-21

Mrs Natalie Boot

Business Support Administrator