

Diversey Europe BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 21 October 2020

Product Name	CLEARKLENS IPA VH1				
Product Code	7513400				
Batch Number	FMP20274 50737				
Production Date	30/09/2020				
Expiration Date	EXP 30/09/2022				

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	T	nits - Upper	Results
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.872	0.883	0.876

0.1.1.10.00	Name:	Angelika Partynska Justyna Staro	
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector	

This document being issued electronically does not bear a signature

COA Template	Version: 02	Date of issuing: November 24 th 2017
--------------	-------------	---



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 1

10/12/2020 12:16:33 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2751385 P.O.# P036537 Sales Order # 2605204

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/11/2020 07:29:58 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	329 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20269	WO50736/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20269	WO50736/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	608 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20274	WO50737/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	940 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.3 KGY	Pass	IWEATHERAL IAN WEATHERALL	10/11/2020 17:28:05 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.8 KGY	Pass	IWEATHERAL IAN WEATHERALL	10/11/2020 17:28:26 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 10/12/2020 12:13:52 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

10/19/2020 8:04:46 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2754150 P.O.# P036578 Sales Order # 2608211

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 10/18/2020 06:25:22 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	97	вх	7513400	Diversey Clearklens IPA 900 mL	FMP20274	WO50737/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20274	WO50737/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	298	BX	7513400	Diversey Clearklens IPA 900 mL	FMP20279	WO50203/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
104.000	2	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20279	WO50203/DEV001152
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	400	вх	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.8 KGY	Pass	BCZAPLA	10/18/2020 16:48:28 GMT
		Reason Code Test				BARTOSZ CZAPLA	
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.1 KGY	Pass	BCZAPLA	10/18/2020 16:48:42 GMT
		Reason Code Test				BARTOSZ CZAPLA	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 10/19/2020 08:02:11 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20274 50737

Lucideon Sample Number: UK204309-30311

Lucideon Report Number: UK204309-30311/ETEP

Issue Number:

Date Logged:

22-Oct-2020

Order Number:

PO 36652

Date Reported:

02-Dec-2020

Date(s) of Test(s):

12-Nov-2020 to 12-Nov-2020

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:


The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

Refer to laboratory investigation report (LIR) 0640.

End of Test Report



Mrs Andrea Saunders

Business Support Assistant

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20274 50737

Lucideon Sample Number: UK204309-30310

Lucideon Report Number: UK204309-30310/MFEP

Issue Number:

1

Date Logged:

22-Oct-2020

Order Number:

PO 36652

Date Reported:

27-Nov-2020

Date(s) of Test(s):

12-Nov-2020 to 26-Nov-2020

Sterility Testing Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report



Mrs Andrea Saunders

Business Support Assistant

Lucideon Limited

Stoke-on-Trent

Staffordshire ST4 7LQ