

Diversey Europe BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tei. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 16 December 2020

Product Name	CLEARKLENS IPA VH1				
Product Code	7513400				
Batch Number	FMP20328 50207				
Production Date	23/11/2020				
Expiration Date	EXP 23/11/2022				

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Li	mits	Results	
		Lower	- Upper		
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid	
Specific Gravity (20°C)	DM004	0.872	0.883	0.875	

On behalf of Diversey site Quality Manager	Name:	Urszula Haraburda Angelika Partynska
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing: November 24th 2017
--------------	--------------	-------------------------------------



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

12/14/2020 10:00:11 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2789030 P.O.# P036952 Sales Order # 2650219

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 12/10/2020 21:45:34 GMT

so UOM **Customer Item Number Customer Item Description** Customer Qty Customer Line # Lot Number Load Number 101.000 960 ВХ 7513400 Diversey Clearklens IPA 900 mL FMP20328 WO50207/DEV001152 Dose Map Ref Pos: 10E1, AF Min = 0.71 302 0030 вх 960 Total

Quality Test Summary

					,	Signed By	
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.6 KGY	Pass	SMITCHELL STEPHEN MITCHELL	12/11/2020 18:13:31 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	42.1 KGY	Pass	SMITCHELL STEPHEN MITCHELL	12/11/2020 18:13:47 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 12/14/2020 09:57:15 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

12/14/2020 10:02:43 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2789036 P.O.# P036955 Sales Order # 2650225

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 12/11/2020 05:10:05 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	387 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20328	WO50207/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP20328	WO50207/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	550 BX	7513400	Diversey Clearklens IPA 900 mL	FMP20331	WO50916/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	940 BX	Total			

Quality Test Summary

				or ourman	,	Signed By	
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.0 KGY	Pass	SWARD SIMON WARD	12/12/2020 12:35:02 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.0 KGY	Pass	SWARD SIMON WARD	12/12/2020 12:35:28 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 12/14/2020 09:59:48 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20328 50207

Lucideon Sample Number: UK205168-36178

Lucideon Report Number: UK205168-36178/ETEP

Issue Number:

Date Logged:

17-Dec-2020

Order Number:

PO 37052

Date Reported:

22-Jan-2021

Date(s) of Test(s):

21-Jan-2021 to 21-Jan-2021

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

15-10t-25 100821

Mrs Natalie Boot

Business Support Administrator

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20328 50207

Lucideon Sample Number: UK205168-36177

Lucideon Report Number: UK205168-36177/MFEP

Issue Number:

Date Logged:

17-Dec-2020

Order Number:

PO 37052

Date Reported:

22-Jan-2021

Date(s) of Test(s):

08-Jan-2021 to 22-Jan-2021

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

JCB001 22-J01-21

Mrs Natalie Boot

Business Support Administrator

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested.