

## CERTIFICATE OF ANALYSIS

Date issue: 14/09/2021


<b>Product Name</b>	DI ClearKlens IPA VH1 S 6x0.9L
<b>Product Code</b>	<input checked="" type="checkbox"/> 101105062/ <input type="checkbox"/> 101105063/ <input type="checkbox"/> 101105064
<b>Batch Number</b>	FMP21334; 53410
<b>Production date</b>	30/11/2021
<b>Expiration Date</b>	EXP 30/11/2023

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Specification	Unit	Results
Appearance	Visual	Colourless Clear Liquid	N/A	Colourless Clear Liquid
Specific Gravity (20°C)	DM 004	0.872 - 0.883	g/cm <sup>3</sup>	0.875

Control of the Sterility of the product:

Test	Test Method	Specification	Conformity	Reference number
Sterilization	Gamma Irradiation	25.0 – 45.0 (kGy)	YES	53410
Sterility certificate	Sterility by Membrane Filtration EP 9.0 2.6.1, USP 42<71> & JP XVII 4.06	No Growth	YES	UK 22116 - 768
2.6.14 Bacterial Endotoxins	Method C – Turbidimetric Kinetic Method (EP/USP Criteria)	<0.2 EU/mL	YES	UK 22116 - 769

	Date of released On behalf of Diversey's Manufacturing location: 31/01/2022	by : Olga Kirchner  Position: QA Admin
---	--	--

The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the SDS.

\*\*\*End of certificate of analysis\*\*\*

R55480102

 12/13/2021 13:55:37 GMT  
 Page 1 of 1

<b>Customer Name:</b>	FLEXIBLE MED. GBP UK LANCASTER	<b>Processing Facility:</b>	Markham Vale	<b>Work Order #</b>	3018970
<b>P.O.#</b>	401439			<b>Sales Order #</b>	2928053
25 - 45 kGy		Gamma Irradiation	Irradiation Date/Time:	12/11/2021 05:08:13 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	289	BX	101105062	Diversey Clearklens IPA 900ml	FMP21322	53409/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	3	BX	101105062_SAMPLE	Diversey Clearklens IPA 900ml	FMP21322	53409/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	658	BX	101105062	Diversey Clearklens IPA 900ml	FMP21334	53410/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
	950	BX	Total			

### Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.1 KGY	Pass	TKVARDA	12/11/2021 17:36:33 GMT
		Reason Code Test				TOMAS KVARDA	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	38.8 KGY	Pass	TKVARDA	12/11/2021 17:36:56 GMT
		Reason Code Test				TOMAS KVARDA	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: STEVEN WHELAN  
 Reason: Work Order Completions

Date: 12/13/2021 13:52:32 GMT

R55480102

 12/20/2021 16:10:00 GMT  
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	3024793
P.O.#	401500			Sales Order #	2934514
	25 - 45 kGy	Gamma Irradiation		Irradiation Date/Time:	12/17/2021 21:17:31 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	176	BX	100862174	Diversey Clearklens DE 900 mL	FMP21343	53418/STR001910
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	771	BX	101105062	Diversey Clearklens IPA 900ml	FMP21334	53410/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	3	BX	101105062_SAMPLE	Diversey Clearklens IPA 900ml	FMP21334	53410/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
	950	BX	Total			

### Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.1 KGY	Pass	JYATES	12/19/2021 06:37:19 GMT
		Reason Code Test				JOHN YATES	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.0 KGY	Pass	JYATES	12/19/2021 06:37:41 GMT
		Reason Code Test				JOHN YATES	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS  
 Reason: Work Order Completions

Date: 12/20/2021 16:07:48 GMT

# PHARMACEUTICAL ANALYSIS REPORT

**LUCIDEON**

insight creating advantage

## Flexible Medical Packaging

Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

FAO: Mrs.Olga Kirchner

**Report of Tests on:** Clear Klens IPA

**Sample Description:** Sample Code: FMP21334 53410

**Lucideon Sample Number:** UK22116-768

**Lucideon Report Number:** UK22116-768/MFEP

**Issue Number:** 1

**Date Logged:** 11-Jan-2022

**Order Number:** PO401662

**Date Reported:** 31-Jan-2022

**Date(s) of Test(s):** 15-Jan-2022 to 29-Jan-2022

### Sterility Testing

Membrane Filtration EP

## Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report**

NcBOOT 31-JAN-22

Mrs Natalie Boot

**Senior Business Support Administrator**

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited  
Queens Road, Penkhull  
Stoke-on-Trent  
Staffordshire ST4 7LQ

T +44 (0)1782 764428  
enquiries@lucideon.com  
www.lucideon.com

# PHARMACEUTICAL ANALYSIS REPORT

# LUCIDEON

insight creating advantage

## Flexible Medical Packaging

Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

FAO: Mrs.Olga Kirchner

**Report of Tests on:** Clear Klens IPA

**Sample Description:** Sample Code: FMP21334 53410

**Lucideon Sample Number:** UK22116-769

**Lucideon Report Number:** UK22116-769/ETEP

**Issue Number:** 1

**Date Logged:** 11-Jan-2022

**Order Number:** PO401662

**Date Reported:** 31-Jan-2022

**Date(s) of Test(s):** 28-Jan-2022 to 28-Jan-2022

### Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

### Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2000 EU/ml)

Tests carried out in accordance with cGMP

**End of Test Report**

Natalie 31-Jan-22

Mrs Natalie Boot

**Senior Business Support Administrator**

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited  
Queens Road, Penkhull  
Stoke-on-Trent  
Staffordshire ST4 7LQ

T +44 (0)1782 764428  
enquiries@lucideon.com  
www.lucideon.com