

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

25 April 2022

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26 first off		
Product Code	7514852		
Batch Number	FMP22115	53046	
Production Date	25/04/2022		
Expiration Date	EXP 25/04/2024		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	9.0	12.5	12.3
Specific Gravity (20°C)	DM004	1.004	1.020	1.004
Potential Chlorine Dioxide (ppm		100	180	148.5

On behalf of Diversey site	Name:	Justyna Staron	Angelika Partynska
Quality Manger	Position	Quality Control Inspecto	

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing: November 24th 2017



Diversey Europe BV Maarssenbroeksedlijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 25/04/2022

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26 - first off		
Product Code	7514852		
Batch Number	FMP22115 53046		
Production Date	25/04/2022		
Expiration Date	EXP 25/04/2024		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method		nits - Upper	Results
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
pH (neat solution)	DM001	1.5	2.5	1.8
Specific Gravity (20°C)	DM004	1.010	1.030	1.014

On bahalf of Diagona in	Name:	J.Staron , A.Partynska
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version: 02	Date of issuing: November 24th 2017

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID 2173-27515A

Product CodeLot NumberQuantityUOMBI-SP B3 Components DV585753050, inc 1 sample box5Case

Validation Reference Number: 5857

Specified Dose Range (kGy): 25.0 - 40.0 Calculated Min Dose (kGy): 29.5 Reference Dose Range (kGy): 26.1 - 35.8 Calculated Max Dose (kGy): 36.0

PO Number: 401161

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Gamma Process Run Approval authorized by STERIS

Date/Time E-Signed: 16-Apr-2022 12:10 AM

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK Limited, a STERIS Company Brunel Close Drayton Fields Industrial Estate Daventry Northants NN11 8RB

Phone: + 44(0) 1327 706 111

Items irradiated under WO53050 will be used in finished batch Fmp22115 53046.

Olga Kirchner

Document ID: 109037

N/A Last Revised in Rel 2.0.0.0 Rel Date: 13-Aug-2018 Page 1 of 1

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Base

Sample Description:

Sample Code: FMP22115 53046

Lucideon Sample Number: UK221760-11176

Lucideon Report Number: UK221760-11176/MFEP

Issue Number:

Date Logged:

28-Apr-2022

Order Number:

402461

Date Reported:

26-May-2022

Date(s) of Test(s):

12-May-2022 to 26-May-2022

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

Mrs Natalie Boot

Senior Business Support Administrator

Page 1 of 1

SOMONSS

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Activator

Sample Description:

Sample Code: FMP22115 53046

Lucideon Sample Number: UK221760-11175

Lucideon Report Number: UK221760-11175/MFEP

Issue Number:

Date Logged:

28-Apr-2022

Order Number:

402461

Date Reported:

26-May-2022

Date(s) of Test(s):

12-May-2022 to 26-May-2022

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

JOBOOF 26 MOY 22

Mrs Natalie Boot

Senior Business Support Administrator

Page 1 of 1