



Diversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427
www.Diversey.com

Certificate of Analysis

Date 16/05/2022

Article
SKU 100934589 ClearKlens oxifast 1L sterile

Page 1/1

Batch: SEP/1581341/22/060
Expiry date: 2024-03
Manufacturing date: 01/03/2022

Characteristic Method	Unit	Value	lower limit	upper limit	conformity
Peracetic acid MAN LCQ 011	%	0,09	0,07	0,10	Yes
Hydrogen peroxide MAN LCQ 011	%	3,04	2,90	3,40	Yes
Loss TAO 16h at 96°C MAN LCQ 009	%	0,45	0,00	2,50	Yes
Sterility certificate European Pharmacopeia 2.6.1					Yes

released on : 17/05/2022

by : PUGAUD

Comments:

Sterility certificate n° 220429-0210-001

The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the MSDS.

End of certificate of analysis