

CERTIFICATE OF ANALYSIS

Date issue: 14/09/2021


Product Name	DI ClearKlens IPA VH1 S 6x0.9L
Product Code	<input checked="" type="checkbox"/> 101105062/ <input type="checkbox"/> 101105063/ <input type="checkbox"/> 101105064
Batch Number	FMP22166; 55050
Production date	15/06/2022
Expiration Date	EXP 15/06/2024

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Specification	Unit	Results
Appearance	Visual	Colourless Clear Liquid	N/A	Colourless Clear Liquid
Specific Gravity (20°C)	DM 004	0.872 - 0.883	g/cm ³	0.875

Control of the Sterility of the product:

Test	Test Method	Specification	Conformity	Reference number
Sterilization	Gamma Irradiation	25.0 – 45.0 (kGy)	YES	55050
Sterility certificate	Sterility by Membrane Filtration EP 9.0 2.6.1, USP 42<71> & JP XVII 4.06	No Growth	YES	UK222786-17787
2.6.14 Bacterial Endotoxins	Method C – Turbidimetric Kinetic Method (EP/USP Criteria)	<0.2 EU/mL	YES	UK222786-17788

	Date of released On behalf of Diversey's Manufacturing location: 02/08/2022	by : Olga Kirchner Position: QA Admin
---	--	--

The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the SDS.

End of certificate of analysis

R55480102

06/27/2022 8:56:45 GMT
Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	3144477
P.O.#	402786			Sales Order #	3073847
	25 - 45 kGy	Gamma Irradiation	Irradiation Date/Time:	06/26/2022 03:22:15 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	425	BX	101105062	Diversey Clearklens IPA 900ml	FMP22166	55050/STR002199
			Dose Map	RF: 13.5A3, AF Min: 0.717		
	425	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By-----	
						User	Date /Time
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.2 KGY	Pass	JBARRATT	06/26/2022 16:13:08 GMT
		Reason Code Test				JAMIE BARRATT	
450.00	Minimum Dose	25.0 kGy	45.0 kGy	28.1 KGY	Pass	JBARRATT	06/26/2022 16:13:26 GMT
		Reason Code Test				JAMIE BARRATT	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
Reason: Work Order Completions

Date: 06/27/2022 08:56:34 GMT

R55480102

 07/04/2022 8:31:08 GMT
 Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	3146781
P.O.#	402840			Sales Order #	3076635
	25 - 45 kGy	Gamma Irradiation		Irradiation Date/Time:	07/02/2022 01:56:28 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	273	BX	101105062	Diversey Clearklens IPA 900ml	FMP22166	55050/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	3	BX	101105062_SAMPLE	Diversey Clearklens IPA 900ml	FMP22166	55050/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	696	BX	101105062	Diversey Clearklens IPA 900ml	FMP22172	55302/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
104.000	3	BX	101105062_SAMPLE	Diversey Clearklens IPA 900ml	FMP22172	55302/STR002199
	Dose Map		302_0112	RF: 13.5A3, AF Min: 0.717		
	975	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.9 KGY	Pass	PSAXTON	07/03/2022 05:12:19 GMT
		Reason Code Test				PAUL SAXTON	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.6 KGY	Pass	PSAXTON	07/03/2022 05:12:38 GMT
		Reason Code Test				PAUL SAXTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: CLAIRE CICALA
 Reason: Work Order Completions

Date: 07/04/2022 08:26:57 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: ClearKlens IPA

Sample Description: Sample Code: FMP22166 55050

Lucideon Sample Number: UK222786-17788

Lucideon Report Number: UK222786-17788/ETEP **Issue Number:** 1

Date Logged: 08-Jul-2022 **Order Number:** 402961

Date Reported: 19-Jul-2022 **Date(s) of Test(s):** 15-Jul-2022 to 15-Jul-2022

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass, <0.2 EU/ml

Tests carried out in accordance with cGMP

End of Test Report

NCBOT 19-JUL-22

Mrs Natalie Boot

Senior Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: ClearKlens IPA

Sample Description: Sample Code: FMP22166 55050

Lucideon Sample Number: UK222786-17787

Lucideon Report Number: UK222786-17787/MFEP **Issue Number:** 1

Date Logged: 08-Jul-2022 **Order Number:** 402961

Date Reported: 02-Aug-2022 **Date(s) of Test(s):** 19-Jul-2022 to 02-Aug-2022

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

 02 AUG 22

Charlotte Metcalfe

Client Technical Support

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com